

RAD-030, RADIATION SAFETY MANUAL**TABLE OF CONTENTS**

TABLE OF CONTENTS	1
LIST OF TABLES	7
INTRODUCTION	8
1. PURPOSE	8
2. SCOPE	8
3. REFERENCES.....	8
3.1. State of Georgia Rules and Regulations on X-Ray, Chapter 290-5-22.....	8
3.2. State of Georgia Rules and Regulations on Radioactive Material, Chapter 391-3-17.....	8
3.3. FDA 21 CFR part 361.1, Radioactive Drugs for Certain Research Uses	8
3.4. OSHA 29 CFR 1910.1096 Ionizing Radiation	8
3.5. Emory University Broad Scope License, GA 153-1	8
4. RESPONSIBILITIES FOR USING RADIATION AT EMORY	8
4.1. General Policy	8
4.2. ALARA Principle	9
4.3. Authorized Users of Radioactive Material – Medical Treatment and Clinical Research	9
4.4. Radiation Permit Holders– non-Human Research	10
4.5. Radiation Workers	10
4.6. Authorized Medical Physicists (AMP).....	11
4.7. Nuclear Medicine Technologists (NMTs)	11
4.8. Authorized Nuclear Pharmacists (ANP).....	11
4.9. X-Ray Users and Operators of Radiation- Producing Equipment	11
5. EMORY UNIVERSITY RESPONSIBILITIES.....	11
5.1. Emory University Management / Executive Oversight	11
5.2. Radiation Control Council.....	12
5.3. Radiation Safety Committee 1	12
5.4. Radiation Safety Committee 2.....	13
5.5. Radiation Safety Committee 3.....	13
5.6. Radioactive Drug Research Committee (RDRC)	13
5.7. Radiation Safety Officer	13
5.8. Deputy Radiation Safety Officer	15
5.9. Directors, Supervisors, and Managers	15
6. ACCESS TO RADIATION SAFETY PROGRAM DOCUMENTS	15
RADIATION EXPOSURE LIMITS & MONITORING	16
7. POLICY	16
8. RADIATION EXPOSURE.....	16
8.1. How Is My Radiation Exposure Measured?	16
8.2. How Much Radiation Can I Receive?	16
8.3. Being Notified of High Exposure/ ALARA Levels.....	16
8.4. Non-Radiation Worker’s Radiation Exposure	17
8.5. How Do I Know What My Radiation Exposure Is?	17
8.6. Who Is Required To Wear Badges?.....	17
8.7. What If I’m Not ‘Required’ to Wear a Badge, But I Think I’m Being Exposed to Radiation?.....	18
8.8. How Do I Request A Badge?.....	18
8.9. Rules for Wearing Badges	19
8.10. When And Where Do I Return Badges?.....	19

RAD-030, RADIATION SAFETY MANUAL

8.11.	Lost Badges	19
8.12.	What If I'm Pregnant?	20
TRAINING PROGRAM.....		21
9.	TRAINING REQUIREMENTS FOR ANCILLARY WORKERS.....	21
9.1.	Training Content.....	21
9.2.	How Do I Get Training?	21
9.3.	Retraining Frequency.....	21
10.	TRAINING REQUIREMENTS FOR LABORATORY WORKERS using radioactive material	21
10.1.	Training Content.....	21
10.2.	How Do I Get Training?	22
10.3.	Retraining Frequency.....	22
11.	TRAINING REQUIREMENTS FOR HEALTHCARE WORKERS using radioactive material	23
11.1.	Training Content.....	23
11.2.	How Do I Get Training?	23
11.3.	Retraining Frequency.....	23
12.	TRAINING REQUIREMENTS FOR X-RAY MACHINE OPERATORS.....	23
12.1.	Training Content.....	24
12.2.	How Do I Get Training?	24
12.3.	Retraining Frequency.....	24
AUTHORIZATION TO USE RADIATION - HOW TO OBTAIN OR AMEND		25
13.	FOR NON-HUMAN USE OF RADIOACTIVE MATERIAL.....	25
13.1.	Application	25
13.2.	Amendments to Authorizations	25
13.3.	Renewals.....	26
13.4.	Active and Inactive Labs	26
13.5.	Suspension	27
14.	FOR HUMAN USE OF RADIOACTIVE MATERIALS.....	27
14.1.	Application	27
14.2.	Amendment	27
14.3.	Renewal	28
14.4.	Suspension	28
15.	CLOSING/TERMINATING A LAB.....	28
16.	FOR MACHINE-PRODUCED RADIATION IN RESEARCH.....	29
16.1.	Human Use and Clinical Trials.....	29
16.2.	Non-Human Use	30
17.	FOR UNESCORTED ACCESS TO IRRADIATORS.....	30
17.1.	How Do I Get Access To An Irradiator?	30
17.2.	But I Need Access Now!	30
SAFE USE OF RADIOACTIVE MATERIAL.....		31
18.	GENERAL RULES FOR HANDLING RADIOACTIVE MATERIAL	31
18.1.	Personal Protective Equipment (PPE)	31
18.2.	Contamination Control	31
18.3.	Exposure Control	31
18.4.	Additional Precautions in Shared or Multi-User Labs.....	31
19.	LABORATORY REQUIREMENTS.....	32
19.1.	Required Lab Features.....	32

RAD-030, RADIATION SAFETY MANUAL

19.2.	Additional Features Dependent on Usage.....	32
20.	GENERAL RULES FOR STORAGE & SECURITY OF RADIOACTIVE MATERIAL	33
20.1.	Self-Contained laboratory.....	33
20.2.	Shared or Multi-User Labs	33
21.	CLINICAL AREAS	33
22.	RADIOACTIVE VOLATILES AND GASES	33
22.1.	Special Requirements for Users of Radioiodine-125 and -131.....	33
23.	SIGNS & POSTINGS	34
23.1.	What Signs Do I Need To Post And Where?.....	34
23.2.	Posting Exceptions.....	34
23.3.	Labeling Requirements.....	35
24.	WIPE TESTS AND GEIGER SURVEYS	35
24.1.	Geiger Surveys	35
24.2.	Wipe Tests	35
24.3.	When to Perform	35
24.4.	How to Perform	36
24.5.	Required Frequency.....	36
24.6.	Action Levels – What Are They & What If They Are Exceeded	37
25.	SEALED SOURCES.....	38
25.1.	Leak-Tests & Inventory of Sealed Sources.....	38
	ACCIDENTS / INCIDENTS / EMERGENCIES	40
26.	PERSONAL CONTAMINATION	40
26.1.	What If I Am Contaminated?.....	40
26.2.	What If I Am Injured?	40
26.3.	How Do I Reach Medical Help?.....	40
26.4.	How to Record Incident and Exposures?.....	41
27.	WHAT IF I HAVE A SPILL?.....	41
27.1.	Minor Spills of Liquids and Solids	41
27.2.	Major Spills of Liquids and Solids	42
28.	OTHER INCIDENTS	42
28.1.	Incidents Involving Dust, Mist, Fumes, Organic Vapors or Gases.....	42
28.2.	Minor Fires	43
28.3.	Major Fire, Explosion or Major Emergencies	43
28.4.	Release into Environment.....	44
28.5.	What If Radioactive Material Is Missing?	44
	PROCUREMENT	45
29.	PROCUREMENT OF RADIOACTIVITY.....	45
29.1.	How to Order Radioactive Material for Research Laboratories	45
29.2.	How to Transfer from Other Universities/ Vendor Replacements.....	45
29.3.	How to Transfer Within the University to Another Lab	45
29.4.	How to Order Sealed Sources.....	46
29.5.	Ordering Radioactive Material for Clinical Departments	46
	RECEIVING RADIOACTIVE MATERIAL	47
30.	RECEIVING PACKAGES OF RADIOACTIVITY.....	47
30.1.	Procedure for Receiving Packages.....	47
30.2.	Who Can Receive Radioactive Material Directly Without Going Through Radiation Safety?.....	47

RAD-030, RADIATION SAFETY MANUAL

MAINTENANCE OF RADIOACTIVE MATERIAL INVENTORY.....	48
31. RADIOACTIVE MATERIAL INVENTORY	48
31.1. What Radioactive Material Am I Supposed to Track?	48
31.2. How Do I Learn To Use EHSAssistant?.....	48
31.3. What If I Made a Mistake?	48
TRANSPORTING RADIOACTIVE MATERIAL	49
32. TRANSPORTATION OF RADIOACTIVITY FROM EMORY	49
32.1. Can I Drive Radioactive Material Around Myself?.....	49
32.2. How Do I Ship Radioactive Material to Somewhere Else?	49
WASTE MANAGEMENT	50
33. LABORATORY AREAS	50
33.1. How Do I Collect Radioactive Waste Safely?	50
33.2. How Should I Separate My Radioactive Waste?	50
33.3. Can I Pour Liquid Waste Down The Drain?.....	51
33.4. How Do I Tell Radiation Safety My Waste Is Ready For Pick-Up?	52
33.5. When Will My Waste Be Picked-Up?.....	52
33.6. Can I Let My Waste Decay Away In My Lab?	52
34. CLINICAL AREAS	52
USE OF RADIOACTIVE MATERIAL IN ANIMALS.....	53
35. SAFE USE OF RADIOACTIVITY IN VIVO	53
35.1. General Guidance	53
35.2. Standard Procedures	53
35.3. Standard Protocols for Care of Animals Containing Radioactivity	54
36. GUIDELINES FOR SELECTING PROPER ANIMAL CARE PROTOCOL	55
36.1. Blue Protocol – Used with any animal that does not require care by radiation workers or isolation.....	55
36.2. Yellow Protocol - Used with any animal requiring care by radiation workers, but not requiring isolation.....	56
36.3. Red Protocol - Used for any animal requiring isolation.	56
INSTRUMENTATION.....	57
37. GENERAL INFORMATION	57
37.1. What Instruments Do I Need To Work With Radioactivity?.....	57
37.2. How Do I Obtain A GM Survey Meter?.....	57
37.3. What Are the Requirements Of GM Survey Meter and Survey Instruments?.....	57
37.4. How Do I Get My GM Survey Meter Calibrated?.....	57
37.5. How Long Does Calibration Take?	58
37.6. How Does the Meter Loaner Program Work?	58
37.7. How Do I Get My GM Survey Meter Repaired?.....	58
37.8. QC Requirements for Gamma Counters/Well Counters/Liquid Scintillation Counters/Thyroid Probes.....	58
37.9. What Dose Calibrator QC is Required?.....	58
37.10. How Do I Dispose Of Old Instruments Such As A Liquid Scintillation Counter?	59
37.11. How Do I Order Instrument QC Sources?	59
37.12. What Should I Do If My Survey or Counting Instruments Fail?	59
37.13. Other Radiation Equipment	59
USE OF RADIOACTIVE MATERIAL IN HUMANS.....	60
38. WHO IS AUTHORIZED TO ADMINISTER RADIOACTIVE MATERIAL TO HUMANS?	60
38.1. Can Any Physician Read Nuclear Studies or Prescribe Radioactive Material for Human Use?	60
38.2. How Do I Become an Authorized User?	60

RAD-030, RADIATION SAFETY MANUAL

38.3.	Who Else Is Authorized To Administer Radioactive Material To Patients or Research Subjects?	61
38.4.	New Nuclear Medicine Technologists.....	61
38.5.	New Authorized Medical Physicists.....	61
39.	GENERAL REQUIREMENTS FOR CLINICAL USE OF RADIOACTIVE MATERIAL	62
39.1.	Use of Protective Equipment	62
39.2.	What are the Rules for Using Gases, Aerosols and Volatiles?	62
39.3.	Rules for Safe Dose Administration	62
39.4.	Special Consideration for the Pregnant or Breastfeeding Patient	63
39.5.	What Information Must Dose Administration Records Contain?.....	63
40.	GUIDELINES FOR WRITTEN DIRECTIVES (PRESCRIPTIONS).....	64
40.1.	Which Administrations Require a Written Directive?.....	64
40.2.	What Information Must Be On the Written Directive?.....	64
40.3.	Written Directive Procedures.....	64
41.	BRACHYTHERAPY RULES	65
41.1.	Technical Requirements	65
41.2.	Determining Brachytherapy Source Activity:.....	66
41.3.	Acceptance Testing.....	66
41.4.	Brachytherapy Procedures	66
41.5.	Specific Nursing precautions:.....	66
41.6.	Brachytherapy Inventory	66
42.	HIGH DOSE RATE (HDR) AFTERLOADER RULES.....	67
42.1.	Installation, Maintenance, Adjustment, & Repair.....	67
42.2.	HDR Procedures	67
42.3.	Safety Procedures	67
42.4.	Required Training.....	68
42.5.	Full-Calibration Measurements Requirements	68
42.6.	Spot Checks	68
43.	PATIENT CARE	68
43.1.	Release of Patients Containing Radioactive Material.....	68
44.	NURSING CARE OF PATIENTS CONTAINING THERAPEUTIC AMOUNTS OF RADIOACTIVE MATERIAL	
	69	
44.1.	Instructions for Nurses:.....	69
44.2.	Inpatient Radiation Controls.....	69
45.	REPORTING PATIENT MEDICAL EVENTS	70
45.1.	Incidents to Report.....	70
	AUTHORIZATION INSPECTIONS	71
46.	LABORATORY AUDITS	71
46.1.	Inspection Procedure	71
46.2.	Keys to a Successful Inspection:	71
46.3.	Reports.....	72
46.4.	Repeat Violations.....	72
46.5.	Corrective Actions	72
47.	CLINICAL AUDITS.....	72
47.1.	Inspection Procedure	72
47.2.	Reports.....	72
47.3.	Corrective Actions	72
47.4.	Surveys by Radiation Safety.....	73
48.	RADIATION SAFETY OFFICE AUDIT RECORDS	73



RAD-030, RADIATION SAFETY MANUAL

MACHINE-PRODUCED RADIATION 74

49. REQUIREMENTS FOR POSSESSION OF RADIATION-PRODUCING EQUIPMENT 74

 49.1. Registration 74

 49.2. Posting 74

 49.3. Surveys 74

 49.4. Shielding for X-Ray Machines 75

50. POLICIES FOR STAFF 75

 50.1. General Policies for Safe Use of X-Ray Equipment 75

 50.3. Can I Be In The Room During An X-Ray Procedure? 77

 50.4. Do I Have To Wear A Lead Apron? 77

51. PATIENT SAFETY GUIDELINES 77

 51.1. Protocol Review 77

 51.2. Fluoroscopy records 77

 51.3. Thresholds for follow up: 78

52. USING X-RAYS IN RESEARCH 78

53. NON-HUMAN USE OF X-RAYS 78

APPENDIX A - LABORATORY INSPECTION VIOLATIONS 79

LIST OF TABLES

Table 1, Exposure Limits for Adults	16
Table 2, ALARA Investigational Levels	17
Table 3, Job Functions at risk of exceeding ALARA 1 Level	18
Table 4, Action Level for Ambient Surveys	37
Table 5, Wipe Test Action Levels - Unrestricted Areas	37
Table 6, Wipe Test Action Levels - Restricted Areas	37
Table 7, Radionuclides Commonly Used in Research Laboratories at Emory University	39
Table 8, Radiotoxicity Classification	54
Table 9, Maximum Permitted Daily Activity Excretion	55

RAD-030, RADIATION SAFETY MANUAL**INTRODUCTION****1. PURPOSE**

The Emory University Radiation Safety Manual is written for the purpose of administering the rules and regulations for the Type A Broad Scope License by specifying the requirements that shall be adhered to by researchers and clinicians. Further, this manual defines the level of compliance required of individuals who wish to utilize radiation or radioactive materials in their research, clinical practice and teaching programs at Emory University and associated institutions.

The requirements of this Radiation Safety Manual are authorized by the Radiation Control Council of Emory University.

All radioactive materials used at Emory University under the Type A Broad Scope License are under the jurisdiction of the State of Georgia, Department of Natural Resources, Radioactive Materials Program. A “Type A Specific License of Broad Scope” is a specific license authorizing receipt, acquisition, ownership, possession, use and transfer of any chemical or physical form of the radioactive material specified in the license, but not exceeding quantities specified in the license, for any authorized purpose. The license may authorize any use of approved radioactive material by anyone in accordance with review and approval procedures established by the Radiation Control Council. Therefore, individuals are not named on the license as users of radioactive material nor are radionuclides limited to narrow, specific uses.

2. SCOPE

This manual applies to all Emory University employees, students and volunteers at any of the Emory College, Emory University School of Medicine, Emory University Hospital, The Emory Clinic, Emory University Hospital at Midtown, Emory University Orthopedic and Spine Hospital and Wesley Woods Geriatric Hospital facilities.

3. REFERENCES

- 3.1. *State of Georgia Rules and Regulations on X-Ray, Chapter 290-5-22*
- 3.2. *State of Georgia Rules and Regulations on Radioactive Material, Chapter 391-3-17*
- 3.3. *FDA 21 CFR part 361.1, Radioactive Drugs for Certain Research Uses*
- 3.4. *OSHA 29 CFR 1910.1096 Ionizing Radiation*
- 3.5. *Emory University Broad Scope License, GA 153-1*

4. RESPONSIBILITIES FOR USING RADIATION AT EMORY**4.1. General Policy**

All employees, students and volunteers are required to comply with the rules set forth by this Manual. All uses of radioactive material must be carried out in accordance with the State of Georgia Rules and Regulations for Radioactive Materials, Chapter 391-3-17, this Manual and written radiation safety procedures applicable to specific areas. All uses of radiation producing machines (RPM) must be carried

RAD-030, RADIATION SAFETY MANUAL

out in accordance with State of Georgia Rules and Regulations on X-Ray, Chapter 290-5-22, this Manual, and written radiation safety procedures applicable to specific areas.

- 4.1.1. All employees, students and volunteers are required to report promptly to the Radiation Safety Office any condition which may cause unnecessary exposure to radiation or radioactive material; or may constitute, lead to, or cause Emory to be in violation of the rules and regulations for radioactive materials or conditions of the license.
- 4.1.2. All employees, students and volunteers are required to complete requisite Emory training before use of radioactive material or radiation producing machines.

4.2. ALARA Principle

The fundamental objective of radiation protection is to keep all radiation exposures ALARA (As Low As Reasonably Achievable) consistent with the purpose for which the activity is undertaken. Exposures are maintained ALARA by following the basic principles of radiation protection, optimizing the amount of time of the exposure, distance from the source of the exposure, and use of appropriate shielding and technology.

4.3. Authorized Users of Radioactive Material – Medical Treatment and Clinical Research

NOTE: Authorized Users who fail to comply with these regulations may cause the University to be subject to license revocation and/or other sanctions provided by law including monetary fines.

Authorized Users are physicians who have been authorized by the Radiation Safety Committee 1 to prescribe and direct the use of radioactive material to humans for clinical or research use and to perform the final interpretation of the results of the tests or studies. If you are a physician wishing to be listed as an Authorized User see “Authorizations. Their responsibilities include:

- 4.3.1. The health and safety of anyone using or affected by the use of radioactive materials under his or her direction or supervision;
- 4.3.2. Receiving initial training and ensuring that his/her staff receive appropriate training;
- 4.3.3. Ensuring that only those individuals trained and designated in writing by an Authorized User are permitted to administer radioactive material to patients or human research subjects;
- 4.3.4. Ensuring that his/her employees, staff, and visitors comply with relevant regulations, policies, and procedures;
- 4.3.5. Reviewing the work of the supervised individual(s) and their radioactive material records;
- 4.3.6. Prescribing the radiopharmaceutical dosage to be administered by issuing a written directive or reference to the diagnostic clinical procedures manual;
- 4.3.7. Preparing and administering, or supervising the preparation and administration of radioactive material for medical use in accordance with applicable policies and regulations;
- 4.3.8. Being physically present for the administration of therapeutic doses;
- 4.3.9. Being immediately available to communicate with the supervised user(s);
- 4.3.10. Informing the Radiation Safety Office of any changes to the authorization;

RAD-030, RADIATION SAFETY MANUAL

- 4.3.11. Reporting any medical events involving radioactive material, such as misadministrations, unintended administrations to pregnant women, etc. to the Radiation Safety Officer.
- 4.3.12. Reporting any medical events involving external beam radiation therapy; the total dose delivered differs from the prescribed dose by 20 percent or more, the fractionated dose delivered differs from the prescribed dose, for a single fraction, by 50 percent or more; an administration of a dose to the wrong individual or human research subject, an administration of a dose or dosage delivered by the wrong mode of treatment (photon vs. electron), administration to the wrong site, etc.

4.4. Radiation Permit Holders– non-Human Research

NOTE: Radiation Permit Holders who fail to comply with these regulations may cause the University to be subject to license revocation and/or other sanctions provided by law including monetary fines.

The Radiation Permit Holder (RPH) is the researcher who has been authorized by the Radiation Safety Committee 2 to use radiation or radioactive material in research *in vitro* or *in vivo*. (See the section [“Authorization to Use Radiation”](#) for obtaining Authorization). The RPH’s responsibilities include:

- 4.4.1. Maintaining exposures ALARA to all laboratory personnel, both users of radioactive material and those who do not use radioactive material in their laboratory protocols.
- 4.4.2. Ensuring that only those individuals trained and approved on their Authorization are permitted to use radiation or radioactive material in their laboratories.
- 4.4.3. Ensuring that laboratory personnel using radiation or radioactive material under their supervision are trained and educated in good radiation safety practices which contribute to maintaining exposures ALARA.
- 4.4.4. Reviewing the supervised individual’s use of radiation or radioactive material, providing reinstruction if needed, and reviewing records kept to reflect this use.
- 4.4.5. Cooperating with the Radiation Safety Office during investigations and audits.
- 4.4.6. Reporting promptly to the Radiation Safety Office any condition which may cause unnecessary exposure to radiation or radioactive material; or may constitute, lead to, or cause Emory to be in violation of the rules and regulations for radioactive materials or the radioactive material license.

4.5. Radiation Workers

The responsibilities of the Radiation Worker include:

- 4.5.1. Following the instructions of the supervising RPH/Authorized User.
- 4.5.2. Following written radiation safety procedures or conditions established in the RPH/Clinical Authorization.
- 4.5.3. Maintaining their radiation exposures ALARA.
- 4.5.4. Properly wearing and returning in a timely manner any personnel monitoring badges issued.
- 4.5.5. Wearing appropriate protective clothing and using proper shielding when indicated.
- 4.5.6. Reporting promptly to the Radiation Safety Office any accidents, incidents or condition which may cause unnecessary exposure to radiation or radioactive material; or may constitute, lead to,

RAD-030, RADIATION SAFETY MANUAL

or cause Emory to be in violation of the rules and regulations for radioactive materials or the radioactive material license.

4.6. Authorized Medical Physicists (AMP)

In addition to the responsibilities of the Radiation Worker in the section above, Authorized Medical Physicists' responsibilities include:

- 4.6.1. Performing full calibration measurements on the high-dose-rate afterloader (HDR).
- 4.6.2. Verifying biweekly that Spot-Checks were performed properly.
- 4.6.3. Performing radiation surveys of the HDR after use.
- 4.6.4. Performing output/activity measurements or calculations for any brachytherapy or HDR treatment plans.
- 4.6.5. AMPs must meet regulatory requirements for training and experience prior to performing any of the above responsibilities.

4.7. Nuclear Medicine Technologists (NMTs)

In addition to the responsibilities of the Radiation Worker in the section above, NMT's responsibilities include:

- 4.7.1. Following requirements for procedures involving written directives or clinical procedures manual when preparing or administering radioactive materials to patients or human subjects (see the section "[Human Use of Radioactive Material](#)");
- 4.7.2. Performing and documenting measurements of radiopharmaceuticals for patients.
- 4.7.3. NMTs must meet regulatory requirements for training and experience prior to preparing radioactive material for human use.

4.8. Authorized Nuclear Pharmacists (ANP)

Authorized Nuclear Pharmacists must meet regulatory requirements for training and experience prior to preparing radioactive material for human use, as reviewed and approved by Radiation Safety Committee 1.

4.9. X-Ray Users and Operators of Radiation- Producing Equipment

Operators of radiation-producing equipment must meet regulatory training requirements. Responsibilities of staff working with or around radiation-producing equipment are detailed in the [Machine-Produced Radiation](#) section of this Manual.

5. EMORY UNIVERSITY RESPONSIBILITIES**5.1. Emory University Management / Executive Oversight**

Emory University Management is directly involved with the Radiation Safety Program. Management provides sufficient authority and organizational freedom to the Radiation Safety Officer and the Radiation Control Council to identify radiation safety problems, to initiate recommendations or provide solutions, and to verify implementation of corrective actions. Emory University Management at the level of Vice President is represented on the Radiation Control Council. Members of the Radiation Control

RAD-030, RADIATION SAFETY MANUAL

Council, Committees 1, 2 and 3, and the Radioactive Drug Research Committee (RDRC) are appointed by Emory University administration.

5.2. Radiation Control Council

The Radiation Control Council serves as the general policy-making and internal regulating body for activities at Emory University that involve the use of radiation and radioactive material. The Radiation Control Council delegates authority to the Radiation Safety Office for enforcement of Radiation Safety policies and procedures. The Radiation Control Council meets periodically throughout the year and reviews and discusses matters relating to the use of radiation at Emory. Specific information about the membership and meetings of the Radiation Control Council can be found in [Radiation Safety Committees](#).

The duties and responsibilities of the Radiation Control Council include, but are not limited to:

- 5.2.1. Review the program for maintaining exposures doses ALARA and providing any additional recommendations needed to ensure that exposures are ALARA.
- 5.2.2. Review and discuss personnel dosimetry data at each Council meeting for personnel exposures exceeding ALARA Level 1 and ALARA Level 2.
- 5.2.3. Review reports of new users and new uses of radioactive materials which have been reviewed and approved by the Radiation Safety Committees.
- 5.2.4. Review and discuss any significant incidents including spills, contamination, and misadministrations with respect to cause and subsequent action taken.
- 5.2.5. Review and approve or disapprove policy and procedural changes prior to implementation. With assistance of Radiation Safety Officer, the Council will audit approved changes to determine implementation and compliance, take appropriate action when noncompliance is identified, and analyze cause and corrective actions to prevent recurrence. Documentation of changes will be maintained stating the reason for the change and summarizing the radiation safety matters that were considered prior to approval of the change.
- 5.2.6. Review, discuss, and approve or disapprove the results of the annual report of the Radiation Safety Program.
- 5.2.7. Due to the scope of the radiation protection program at Emory, four subcommittees operate under the control of the Radiation Control Council to review and give initial approval to users and uses of radioactive material, depending on the type of use:
 - 5.2.7.1. Radiation Safety Committee 1 – Human Use of Radioactive Material
 - 5.2.7.2. Radiation Safety Committee 2 – Non-Human Use of Radiation
 - 5.2.7.3. Radiation Safety Committee 3 – Human Use of Machine Produced Radiation in Research
 - 5.2.7.4. Radioactive Drug Research Committee

5.3. Radiation Safety Committee 1

Radiation Safety Committee 1 reviews the human use of radioactive material at Emory. Specific information about the membership and meetings of the Radiation Safety Committee 1 can be found in [Radiation Safety Committees](#).

The duties and responsibilities of the Radiation Safety Committee 1 include:

RAD-030, RADIATION SAFETY MANUAL

- 5.3.1. Approving persons applying to function as Authorized Users, Authorized Medical Physicists, or Nuclear Medicine Technologists;
- 5.3.2. Reviewing and approving the procedures, types and quantities of radioactive materials requested for human research or clinical use;
- 5.3.3. Consider problems brought before it by the Radiation Safety Officer, medical staff members or technical staff and give a prompt ruling in each case.

5.4. Radiation Safety Committee 2

Radiation Safety Committee 2 reviews applications and amendments for the non-human research use of radioactive material and machine-produced radiation with respect to user qualifications, types and quantities of materials requested, and uses of materials requested. Specific information about the membership and meetings of the Radiation Safety Committee 2 can be found in [Radiation Safety Committees](#).

The duties and responsibilities of the Radiation Safety Committee 2 include:

- 5.4.1. Review on the basis of safety and approve or disapprove each proposed method of non-human use of radioactive material.
- 5.4.2. Consider problems brought before it by the Radiation Safety Officer or other interested parties and give a prompt ruling in each case.

5.5. Radiation Safety Committee 3

Radiation Safety Committee 3 reviews applications and amendments for the research use of machine-produced radiation on humans with respect to procedure, exposure to the subject, and risk information provided to the subject. Specific information about the membership and meetings of the Radiation Safety Committee 3 can be found in [Radiation Safety Committees](#).

The duties and responsibilities of the Radiation Safety Committee 3 include:

- 5.5.1. Review on the basis of safety and approve or disapprove each proposed method of machine-produced radiation on humans
- 5.5.2. Recommend or approve policy for the safe use of x-rays in human research;
- 5.5.3. Consider problems brought before it by the Radiation Safety Officer or other interested parties and give a prompt ruling in each case.

5.6. Radioactive Drug Research Committee (RDRC)

The Radioactive Drug Research Committee #40 (RDRC) is charged by the FDA to approve and track research in humans using radioactive drugs generally recognized as safe and effective in order to obtain basic information regarding drug metabolism, human physiology, pathophysiology or biochemistry.

5.7. Radiation Safety Officer

The Radiation Safety Officer is responsible for day-to-day oversight of the Radiation Safety Program. Duties and responsibilities include:

- 5.7.1. Develop, distribute, and implement up-to-date protection procedures in the daily operation of the Emory University radiation protection program;

RAD-030, RADIATION SAFETY MANUAL

- 5.7.2. Ensure that possession, use, and storage of radioactive materials are consistent with the limitations in license GA 153-1 and Georgia Rules and Regulations for Radioactive Materials, Chapter 391-3-17;
- 5.7.3. Ensure that personnel training is conducted and is commensurate with the individual's duties regarding radioactive material;
- 5.7.4. Maintain documentation to demonstrate, by measurement or calculation, that individuals are not likely to receive, in one year, a radiation dose in excess of 10% of the allowable limits; or in the alternative, ensure that personnel monitoring devices are provided;
- 5.7.5. Establish and maintain a personnel monitoring program ensuring that dosimeters are appropriately provided, used, and exchanged, and that records of monitoring are maintained;
- 5.7.6. Ensure that radioactive material is properly secured;
- 5.7.7. Maintain documentation to demonstrate, by measurement or calculation, that the highest total effective dose equivalent to the non-occupationally exposed individual likely to receive the highest dose from the licensed operation does not exceed the annual limit for members of the public;
- 5.7.8. Notify proper authorities of incidents, such as loss or theft of radioactive material, damage to or malfunction of sealed sources, and fire involving radioactive materials;
- 5.7.9. Investigate and report to the State Medical Events and precursor events including cause, appropriate corrective actions identified, and timely corrective actions taken;
- 5.7.10. Participate in and document audit of the radiation protection program at least annually for adherence to ALARA concepts and seek to remedy any deficiencies noted;
- 5.7.11. Identify violations of regulations, license conditions, or program weaknesses and develop, implement and document effective corrective actions;
- 5.7.12. Ensure that radioactive material is transported, or offered for transport, in accordance with all applicable Department of Transportation (DOT) requirements;
- 5.7.13. Dispose of radioactive material properly;
- 5.7.14. Maintain an up-to-date license, and submit amendments and renewal requests in a timely manner;
- 5.7.15. Review quarterly radiation levels in restricted areas and adjacent non-restricted areas as indicated on staff surveys;
- 5.7.16. Review or develop shielding plans for new radiation areas; and,
- 5.7.17. Conduct annual surveys on X-Ray equipment and provide scatter and exposure charts to designated departments.
- 5.7.18. Audit all active radioactive material use and storage facilities quarterly and report any findings to the primary Radiation Permit Holder and the appropriate Radiation Safety Committee.
- 5.7.19. Stop unsafe operations that come to the RSO's attention.

RAD-030, RADIATION SAFETY MANUAL**5.8. Deputy Radiation Safety Officer**

The Deputy Radiation Safety Officer(s) operate under the direction of the Radiation Safety Officer and meet the training and experience criteria for medical Radiation Safety Officers per regulatory requirements and may participate in any role requiring the Radiation Safety Officer's authority.

5.9. Directors, Supervisors, and Managers

- 5.9.1. Directors, supervisors, and managers must have knowledge of the use of radioactive material or radiation producing machines in their areas.
- 5.9.2. They must understand that certain uses of ionizing radiation sources require specific training for the worker that might not be provided by EHSO (for example, certification training for nuclear medicine technologists or radiological technologists).
- 5.9.3. They must be aware that additions, removals, or alterations in the use of such radiation sources may require actions to be carried out by the Radiation Safety Officer for safety or regulatory purposes, and must therefore keep the Radiation Safety Officer informed. Examples of actions that would require informing the Radiation Safety Officer include:
 - 5.9.3.1. Hiring persons to serve as Authorized Users (AUs), Authorized Medical Physicists (AMPs), or Nuclear Medicine Technologists (NMTs);
 - 5.9.3.2. Whenever an Authorized User, Authorized Medical Physicist, or Nuclear Medicine Technologist leaves the Emory system;
 - 5.9.3.3. Whenever a Radiation Permit Holder (RPH) leaves;
 - 5.9.3.4. Procuring new or disposal of old radiation-producing equipment;
 - 5.9.3.5. Relocating or reconstruction of radiation areas that alter or require radiation shielding.
- 5.9.4. Additionally, the Radiation Safety Officer may turn to the director, supervisor, or manager for assistance in the enforcement of the radiation protection program.

6. ACCESS TO RADIATION SAFETY PROGRAM DOCUMENTS

All licenses and documents are available for inspection at the Emory University Environmental Health and Safety Office, 1762 Clifton Road, Room 1200, Atlanta, GA 30322.

RAD-030, RADIATION SAFETY MANUAL

RADIATION EXPOSURE LIMITS & MONITORING

7. POLICY

Refer to [Occupational Exposure and Personnel Monitoring Program](#) for additional details on Emory’s program for measuring, reporting, and investigating worker’s exposure to radiation.

8. RADIATION EXPOSURE

8.1. *How Is My Radiation Exposure Measured?*

Exposure to ionizing radiation is measured using monitoring devices called dosimeters or radiation “badges”. Badges measure external exposure to radiation. After the badges are worn for a specified period of time they are sent for commercial processing. The amount of exposure to the badge is reported in units of millirem (mrem). Badges are worn on the part of the body most likely to receive radiation exposure. More information on badge placement is found in [Rules for Wearing Badges](#).

Internal exposure to radiation, from inhalation, ingestion, or absorption, is unlikely to come from most forms of radioactive material used at Emory. However, some use of radioactive material requires monitoring of internal exposure to radiation. Tests used to determine internal exposure are called “bioassays”. A list of those uses of radioactive material requiring bioassays and the type of bioassay is found in [Occupational Exposure and Personnel Monitoring Program, section 3](#).

8.2. *How Much Radiation Can I Receive?*

If you are an adult radiation worker, then your radiation exposure limits are listed in Table 1 below. Occupational exposure limits for minors are 10% of the corresponding limit for adults. If you exceed any annual limit, then you will not be allowed to work with or around radiation for the rest of the year.

Table 1, Exposure Limits for Adults

Annual Occupational Exposure Limits (Radioactivity)		Quarterly Limits for X-Ray
Whole Body	5 rem (0.05 Sv)	1 ¼ rem/qtr (3 rem/qtr allowed if 5 rem/year not exceeded)
Lens of Eye	15 rem (0.15 Sv)	1 ¼ rem/qtr (3 rem/qtr allowed if 5 rem/year not exceeded)
Skin or any extremity	50 rem (0.50 Sv)	7 ½ rem/qtr
Fetal exposure	0.5 rem (5.0 mSv) during entire pregnancy after declaration	not to exceed 0.05 rem per month, 0.5 rem (5.0 mSv) during entire pregnancy after declaration

8.3. *Being Notified of High Exposure/ ALARA Levels*

In order to identify those workers at most risk of exceeding radiation exposure limits, quarterly investigational levels have been established. These levels are called “ALARA Levels”, named after the basic radiation safety principle to always keep your exposure as low as reasonably achievable. There are two levels for each exposure limit; ALARA level 1 and ALARA level 2 (refer to

Table 2). Radiation Safety will notify any worker who receives an exposure in excess of ALARA 1. If the exposure exceeds ALARA 2, then Radiation Safety will also investigate in order to determine whether or not additional measures can or should be taken to reduce the exposure.

Table 2, ALARA Investigational Levels

QUARTERLY ALARA INVESTIGATIONAL LEVELS		
	LEVEL 1	LEVEL 2
Whole Body	> 125 millirems	> 375 mrem
Lens of Eye	> 375 millirems	> 1125 mrem
Skin /Extremities	> 1250 millirems	> 3750 mrem
Fetal	none	> 50 mrem/month

8.4. *Non-Radiation Worker's Radiation Exposure*

If you're not a radiation worker, then, by definition, your annual exposure from the use of radiation at Emory cannot exceed 100 mrem; exclusive of background, medical procedures or exposures from patients authorized for release or sewer disposal. There are many controls in place to keep radiation exposure within this limit for those who do not work with radiation. These include shielding x-ray rooms; documented daily and weekly surveys for contamination and measurements of radiation levels, establishing restricted areas; establishing safety procedures for receiving and disposing of radioactive material, establishing safety procedures for using x-ray equipment, etc.

8.5. *How Do I Know What My Radiation Exposure Is?*

All exposure reports are reviewed by members of the Radiation Safety Office. As stated previously, if your exposure exceeds an ALARA level, you will be notified by Radiation Safety. Otherwise, your Radiation Exposure Reports are sent to the radiation safety contact in your department. You should make sure that you know where those reports are kept. Some departments require you to initial your exposure report to document that you are aware of your exposure.

Annual reports are also prepared by the Radiation Safety Office for each worker as required by regulation. Distribution is by mail directly to the individual, to the department contact, by electronic means, or a combination thereof.

If you ever have any problems finding your radiation exposure reports, please contact Radiation Safety. The Emory Radiation Safety Office maintains copies of all radiation exposure records.

8.6. *Who Is Required To Wear Badges?*

Regulations for both radioactive material and x-ray machines require that any person who is occupationally exposed to radiation at a level which is likely to exceed ten percent of any regulatory limit must be issued a radiation badge. ALARA Level 1 exposures, seen in [Table 2](#) are set at ten percent of the regulatory limits.

[Table 3](#) lists job functions that are at the greatest risk for exceeding an ALARA 1 Level, and the type of badge(s) they will be assigned:

Table 3, Job Functions at risk of exceeding ALARA 1 Level

If your job function involves...	Then you must wear* a ...
Working with large quantities of beta emitters	ring badge
Operating any type of x-ray equipment	body badge
Performing or assisting fluoroscopy procedures infrequently	collar badge
Performing or assisting fluoroscopy procedures on a daily basis	body and collar badge **
Handling radiopharmaceuticals or brachytherapy sources for patients	body and ring badge
Performing HDR procedures	body badge
Nurses caring for patients containing radioactive material for therapy	body badge
<i>* see also section below “Rules for Wearing Badges” ** see section on Wearing Badges for how radiation exposure from two badges is calculated.</i>	

8.7. What If I’m Not ‘Required’ to Wear a Badge, But I Think I’m Being Exposed to Radiation?

Any worker concerned about his or her exposure to radiation from radioactive materials, x-ray machines, or radioactive patients should consult the Radiation Safety Office. The Radiation Safety Office can determine, by your job responsibilities, if a radiation badge is necessary. If one is not necessary, the Radiation Safety Office can explain why you are not likely to receive enough exposure to require monitoring. For example, if someone at your job was monitored previously, the Radiation Safety Office can evaluate your potential exposure based on those exposure records. If you are still concerned about your exposure, then the Radiation Safety Office can issue you a badge for an evaluation period, (typically six months) or until you don’t wish to be monitored anymore.

8.8. How Do I Request A Badge?

Contact the Radiation Safety Office. You will be asked to complete a form (see below) and receive documented training (see Section -[Training](#).) in the proper use of a badge and basic radiation protection principles.

8.8.1. Forms needed for Badge Requests:

- 8.8.1.1. Healthcare workers must complete the Personnel History Monitoring Form,
- 8.8.1.2. Laboratory workers must complete the Training and Personnel Monitoring Determination Form.

Once you’ve completed the forms and received the training, a badge will be issued to you. All workers issued radiation badges are required to follow those rules and responsibilities stated in this Manual and to cooperate with the Radiation Safety Office in their efforts to maintain exposures ALARA.

RAD-030, RADIATION SAFETY MANUAL**8.9. Rules for Wearing Badges**

- 8.9.1. Control badges assigned to a shipment must be kept in a low radiation background area. These control badges must be returned when the badges are collected for processing.
- 8.9.2. If you are issued a badge, you must wear it whenever you are working with or near sources of radiation exposure.
- 8.9.3. Badges must only be worn at work.
- 8.9.4. Wear the badge on the part of the body closest to the source of radiation.
- 8.9.5. For x-ray machine operators or staff issued a single badge, it must be worn unshielded on the collar or waist area. If a lead apron is worn, it must be worn on the outside of the apron at the collar. Assigned dose to Whole Body (WB) is 0.3 times the collar badge reading.
- 8.9.6. For x-ray machine operators or staff issued two badges, a designated chest badge must be worn on the torso, shielded underneath the lead vest. The designated collar badge must be worn unshielded at the collar. The collar badge will be used to evaluate the eye and skin exposure. The worker's whole body dose (WB) will be calculated from the results of the two badges according to the following formula:
$$WB = (1.5 \times \text{Body Badge}) + (0.04 \times \text{Collar Badge})$$
- 8.9.7. Ring monitors must be worn so that the label is facing the palm of your hand and underneath your glove.
- 8.9.8. Wear a ring badge, if required, during the elution of generators; during the preparation, assay, and injection of radiopharmaceuticals; when performing camera quality control; when holding patients during procedures; and when handling radioactive material.
- 8.9.9. Fetal monitors must be worn at the abdomen, under any protective lead.
- 8.9.10. Badges, when not being worn, must not be stored near any radiation sources.
- 8.9.11. A badge assigned to one individual must never be worn by another individual.

8.10. When And Where Do I Return Badges?

Departments shall establish a Radiation Safety contact that is responsible for exchanging the old and new badges. Laboratories on Emory campus can use Interoffice Mail to send their badges to the Radiation Safety Office. Badges in clinical departments at EUH, Emory Clinic, etc. are typically picked up by Radiation Safety Office personnel.

Please make every effort to return your badges for processing by the 8th of the month following the wear period. Late badges generate unnecessary work effort and expense. The Radiation Safety Office may request disciplinary advice from the supervisors of workers who chronically return badges late.

8.11. Lost Badges

Report all lost badges to the Radiation Safety Office. If you need a replacement, one can be assigned. Radiation Safety staff may ask if you performed more, less or any unusual work with radiation during the wear period of the lost badge so that an exposure amount can be estimated and assigned to your history.

RAD-030, RADIATION SAFETY MANUAL**8.12. *What If I'm Pregnant?***

Employees in the Emory community who work with radiation have the option of notifying and are encouraged to notify the Radiation Safety Officer (RSO) of suspected or confirmed pregnancies as soon as possible, so that the RSO can work with the employee and her supervisor to monitor the radiation exposure levels during the pregnancy and take measures, as appropriate, to maintain exposures as low as reasonably achievable and within regulatory limits.

All radiation workers will be informed of applicable state and federal regulations regarding occupational exposure to the fetus from ionizing radiation during their initial radiation safety training.

As soon as a radiation worker determines that she is pregnant, she should (unless privacy is desired) advise her supervisor and declare her pregnancy in writing to the Radiation Safety Office using the [Form Letter for Declaring Pregnancy](#) to the Radiation Safety Office giving the approximate date of conception. A professional from the Radiation Safety Office will review the past radiation exposure history of the declared pregnant radiation worker and her job function and determine if radiation restrictions should be applied. If so, these restrictions will be discussed with the individual and her supervisor and will be provided to both in writing. A copy of the document "[Guide for Instruction Concerning Prenatal Radiation Exposure](#)" will be given to the declared pregnant radiation worker as required by the State, NRC and OSHA. The employee and supervisor (unless privacy is desired) will complete the [Emory Radiation Safety Consultation](#) form documenting that this information has been given. A radiation worker may, without declaring pregnancy, consult with the Radiation Safety Office concerning issues relating to exposure of an embryo/fetus to radiation in the course of the employee's job.

The Radiation Safety Office will issue a monthly fetal badge for the declared pregnant radiation worker to wear at the waist in addition to her regular badge.

All lead barriers in the university are designed so an individual, if she were behind the barrier for the full 40 hours of a week, would receive less than 10 mrem to the surface of her body and much less to the fetus. NCRP, NRC and the State of Georgia allow the fetus of a declared pregnant radiation worker to receive 500 mrem, sum of internal and external exposure, during the nine months of pregnancy.

It is recommended that pregnant nurses not care for patients containing therapeutic quantities of a radionuclide or brachytherapy sources.

RAD-030, RADIATION SAFETY MANUAL**TRAINING PROGRAM**

EHSO is responsible for ensuring that radiation safety training is provided to all Emory University employees who work with radioactive materials. You must complete training prior to beginning work with radioactive materials or equipment that produces radiation. Instructions are provided to those Emory University employees who work in close proximity to radioactive materials or equipment that produces radiation. Required training is dependent on the job functions performed and falls into four categories: 1) ancillary; 2) laboratories; 3) healthcare; or 4) machine-produced radiation. It is possible that an employee will be trained in more than one category. Training will include regulatory-required content based on the job function, as well as education topics identified by the department in which you work and the radiation safety staff.

Online and classroom courses and self-study are described and can be accessed on the [training page of the EHSO website](#). Training completions shall be documented and retained by EHSO.

9. TRAINING REQUIREMENTS FOR ANCILLARY WORKERS**9.1. Training Content**

Workers who enter radiation use areas to perform their duties with no direct use of radioactive material will receive instruction that includes the following topics:

- Potential radiation hazards in each area where the employees will work
- Posting requirements of areas where radioactive material /radiation is used and/or stored
- Basic radiation protection to include concepts of time, distance, and shielding

9.2. How Do I Get Training?

Training courses and instructions for accessing them are listed on the [training page of the EHSO website](#). Instructions are provided in the new hire orientation, in the EHSO Lab Rat newsletter and on the [Policies and Procedures](#) page of the EHSO website. You also can contact the research radiation safety liaison for your facility: <http://www.ehso.emory.edu/about/contact.html>.

9.3. Retraining Frequency

Retraining will be conducted whenever there is a significant change in duties or regulations.

10. TRAINING REQUIREMENTS FOR LABORATORY WORKERS USING RADIOACTIVE MATERIAL**10.1. Training Content**

Training for laboratory workers will include the following topics:

- Atomic Structure
- Alpha, beta, and gamma radiation
- Radioactivity units
- Radioactive decay
- Biological effects
- Background radiation
- ALARA
- Radiation protection principles

RAD-030, RADIATION SAFETY MANUAL

- Radiation surveys
- Radiation inventory
- Recordkeeping
- Personal protective equipment (PPE)
- Waste disposal
- Occupational dose limits and dosimetry
- Policy on radiation and pregnancy
- Purchase, receipt, and storage of radioactive material
- Radiation instrumentation
- Spill and contamination procedure (non-emergency)
- Emergency response

10.2. How Do I Get Training?

Training courses and instructions for accessing them are listed on the [training page of the EHSO website](#). You also can contact the Radiation Safety health physicist for your facility: <http://www.ehso.emory.edu/about/contact.html>.

10.2.1. You must complete initial training which is covered in the Radiation Safety for Laboratory Research Personnel Course, which is offered as two modules:

10.2.1.1. Module 1 is completed online as a prerequisite to Module 2.

10.2.1.2. Module 2 is completed in a classroom setting and consists of a practicum that is offered monthly. The course is presented by the research radiation safety liaisons who work closely with the research labs.

10.2.2. If you must use radioisotopes before attending the monthly laboratory safety course, you must take the initial, online course – Module 1 and request a personal training session for Module 2 (call 404-727-5922), prior to using the radioisotopes.

10.2.3. For Non-Human uses in vivo, PET imaging research, etc. special courses may be available. See the training courses index.

10.3. Retraining Frequency

Retraining for laboratory workers using radioactive material is required at least every 3 years and is accomplished using one of several mechanisms:

10.3.1. You can repeat the Initial Training.

10.3.2. You will take the online Radiation Safety for Lab Workers Refresher Course, or

10.3.3. You can request an in-service presentation for a laboratory or division:

10.3.3.1. the same general material is presented but is streamlined to fit the needs of the group;

10.3.3.2. more hands-on practice time with survey meters and calculating efficiencies of liquid scintillation counters and dpm can be addressed in greater detail; and

10.3.3.3. Other specific problems or incidents can be discussed in greater depth.

10.3.4. Significant or repeated inspection deficiencies may indicate the need for early retraining as part of corrective actions.

RAD-030, RADIATION SAFETY MANUAL**11. TRAINING REQUIREMENTS FOR HEALTHCARE WORKERS USING RADIOACTIVE MATERIAL****11.1. Training Content**

11.1.1. Training for clinical workers will include the following topics:

- Potential radiation hazards in each area where the employees will work
- Posting requirements of areas where radioactive material /radiation is used and/or stored
- Basic radiation protection to include concepts of time, distance, and shielding
- Basic radiation biology
- Risk estimates, including comparison with other health risks
- Proper use of personnel dosimetry and exposure reporting (when applicable)
- Concept of maintaining exposure ALARA
- Occupational dose limits and their significance
- Dose limits to the embryo/fetus, including instruction on declaration of pregnancy
- Dose to individual members of the public
- Worker's right to be informed of occupational radiation exposure
- Individual's obligation to report unsafe conditions to the Radiation Safety Office
- Applicable regulations, license conditions, information notices, bulletins, etc.
- Location of copies of the applicable regulations, the Radioactive Materials License, and its application are posted or made available for examination
- Access control procedures
- Proper use of radiation shielding, if used
- Emergency procedures.

11.1.2. Additional training is designated for the following specific job function(s)

- Authorized Users;
- Authorized Medical Physicists;
- Nuclear Medicine Technologists;
- Nurses caring for patients containing radioactivity.

11.2. How Do I Get Training?

Training courses and instructions for accessing them are listed on the [training page of the EHSO website](#). You also can contact the research radiation safety liaison for your facility: <http://www.ehso.emory.edu/about/contact.html>.

11.3. Retraining Frequency

Retraining will be conducted whenever there is a significant change in duties or regulations. Refresher training frequency is determined by the terms of the license.

12. TRAINING REQUIREMENTS FOR X-RAY MACHINE OPERATORS

State of Georgia regulations require operators of radiation-producing machinery to a minimum of training. For patient and veterinary use, the requirement is 6 hours of radiation safety training prior to use of x-ray machines on patients. Similar requirements exist for other uses of x-ray – producing machines such as x-ray crystallography machines and DEXA units.

RAD-030, RADIATION SAFETY MANUAL

NOTE: Physicians, Veterinarians, Radiologic Technologists, and Nuclear Medicine Technologists are considered to have satisfied the Georgia certification requirement for machine-produced radiation safety training as part of their state licensing and will not need to complete this training through EHSO.

12.1. Training Content

Training for operators and support staff will include the following topics pertinent to their specific job function:

- Applicable regulations
- Operator responsibilities
- Potential radiation hazards in work areas
- Risk estimates, including comparison with other health risks
- Basic radiation biology
- Steps to minimize exposure to patients & staff
- Patient safety
- Radiation effects on skin
- Pregnant patient precautions
- Recording exposure information
- Reporting high or accidental exposures
- Proper use of personnel dosimetry
- Department-specific work rules
- Procedure risks

12.2. How Do I Get Training?

Training courses and instructions for accessing them are listed on the [training page of the EHSO website](#). You also can contact the research radiation safety liaison for your facility: <http://www.ehso.emory.edu/about/contact.html>.

12.3. Retraining Frequency

Retraining will be conducted whenever there is a significant change in duties or regulations.

RAD-030, RADIATION SAFETY MANUAL**AUTHORIZATION TO USE RADIATION - HOW TO OBTAIN OR AMEND****13. FOR NON-HUMAN USE OF RADIOACTIVE MATERIAL****13.1. Application**

- 13.1.1. The applicant must be a faculty member of a division of Emory University with training and experience commensurate with the intended use.
- 13.1.2. The applicant must complete a [Radiation Safety Committee 2 Application](#). Within this application, you must sign a statement indicating familiarity with the Emory University Radiation Safety Manual and acknowledging your commitment to keeping exposures as low as reasonably achievable.
- 13.1.3. The applicant submits the signed original application to the Radiation Safety Office.
- 13.1.4. After the Radiation Safety Office checks the application for completeness, it is sent to Radiation Safety Committee 2 members where it is reviewed with respect to: training and experience presented by the Radiation Permit Holder in reference to proposed use; facilities and instrumentation available; and proposed techniques of safely using and disposing of radioactive material.
- 13.1.5. Approval is granted by unanimous vote by a convened quorum after resolution of any comments or questions.
- 13.1.6. Upon approval, a Radiation Safety staff member will deliver the Authorization (e.g. Permit), postings, and emergency procedures signs to the laboratory.
- 13.1.7. Laboratory personnel named on the authorization to use radioactive material must receive training prior to their first use of radioactive material.
- 13.1.8. The Radiation Permit Holder (RPH) will evaluate all approved procedures before using radioactive materials to ensure that exposures will be kept ALARA. This may be enhanced through the applications of trial runs.

13.2. Amendments to Authorizations

- 13.2.1. Use the [Amendment to Authorization – Committee 2 form](#) to submit amendments involving the use of radioactive material. Amendments may include adding, removing or changing protocols, adding or removing personnel, or adding or removing labs.
- 13.2.2. An approval of changes of personnel to use isotopes, and addition and deletion of laboratory locations may be made at the discretion of the Radiation Safety Office. Committee members receive a report of these administrative changes at each meeting.
- 13.2.3. Protocol amendments must include, at a minimum: the project title; the radionuclide and chemical form involved; an indication of any need to increase total possession limit or to add a chemical form; a detailed description of the portion(s) of the protocol dealing with the use of the radioactive material; the names and experience of any additional staff members who may be added due to the project; and, IACUC committee approvals where needed.
- 13.2.4. When additional isotopes or procedures are involved, the applicant must demonstrate appropriate knowledge and experience to ensure safety in carrying out the protocol.

RAD-030, RADIATION SAFETY MANUAL

- 13.2.5. The amendment will be circulated through the committee in the same manner as the original application. Approval is granted by unanimous vote by a convened quorum after resolution of any comments or questions.
- 13.2.6. For adding new personnel - all new personnel must receive preliminary training through EHSO before receiving approval.
- 13.2.7. For adding labs – laboratories will undergo a pre-occupancy contamination survey by Radiation Safety personnel;
- 13.2.8. For removing labs – laboratories will undergo a post-occupancy contamination survey by Radiation Safety personnel.
- 13.2.8.1. When adding a lab, the Radiation Safety Office will add the room number(s) to the Authorization and post the lab accordingly. The RPH must then add the lab to the list of areas to be surveyed.
- 13.2.8.2. When removing a lab, the RPH should follow lab decommissioning guidelines. The researcher is responsible for any radioactive material abandoned in a lab.

13.3. Renewals

The RPH must apply for renewal every three (3) years. Authorization renewal forms will be provided to the Radiation Permit Holders prior to expiration.

13.4. Active and Inactive Labs

- 13.4.1. When the RPH has not used radioactive material and has had no radioactive materials inventory or waste for a period of one year, Inactive Status can be requested by amendment. Radiation Safety will review the current standing of the Authorization prior to approval. Authorizations approved for Inactive Status will not be inspected, personnel will not require refresher training, and survey instruments will be stored and will not require annual calibration.
- 13.4.1.1. Reactivation of Inactive Labs: Reactivation will require the lab to submit correspondence to the Radiation Safety Office in the form of a letter or e-mail. Radiation survey instruments must be calibrated and Radiation Safety Training completed by laboratory radiation workers prior to ordering or working with radioactive material. Reactivation may take several weeks to allow for retraining of personnel and calibration of survey instrument(s).
- 13.4.1.2. Laboratories that are placed on Inactive Status will turn in all radiation badges during this period. In the event Active status is requested, badges will be provided after training requirements and a risk assessment has been completed.
- 13.4.1.3. When EHSO staff receives an order or request for transfer of radioactive material to a laboratory on Inactive status, Radiation Safety Office will be notified and the Radiation Permit Holder will be returned to active status once training and survey instrument calibration requirements are satisfied.
- 13.4.2. When the RPH has not used radioactive material and has had no radioactive materials inventory or waste for a period of two years or longer, the Authorization will be terminated and the laboratory decommissioned for radioactive material use.
- 13.4.3. Authorizations must be kept in force. Radiation Permit Holders will be notified when renewals are due. Any changes made during the review will be included in the renewal application.

RAD-030, RADIATION SAFETY MANUAL

Renewal applications received by the expiration date of the authorization will be considered active until Radiation Safety Committee 2 issues a new authorization.

13.4.4. If the Radiation Permit Holder does not submit renewal materials by the expiration date, the authorization will be terminated.

13.5. Suspension

Radiation Safety Committee 2 and the Radiation Safety Office have the right to suspend an authorization, in full or in part, following the discovery of violations or infractions of the authorization. The Committee may reinstate the authorization once the violations have been corrected and an increased surveillance program put into place.

14. FOR HUMAN USE OF RADIOACTIVE MATERIALS**14.1. Application**

14.1.1. Applicants must have faculty status (assistant professor or greater), must be experienced in the use of radioactive materials and must be trained by EHSO prior to approval.

14.1.2. Researchers must apply to Radiation Safety Committee 1 for a Human Use Authorization to use radioactive materials in humans, in either research or clinical applications. [See the [Committee 1 Application for Human RAM Use](#) on the EHSO website].

14.1.3. Applications will require the description of personnel monitoring, protective clothing, shielding and survey procedures designed to properly evaluate exposures and maintain them ALARA.

14.1.4. Submit completed application forms to the Radiation Safety Officer.

14.1.5. Approval is granted by unanimous vote by a convened quorum after resolution of any comments or questions.

14.1.5.1. If a member dissents or questions an application, the matter must be resolved to the member's satisfaction prior to approval.

14.1.5.2. Radiation Safety Committee 1 may require additional conditions under which the use of the material must be conducted.

14.1.6. Once approved, the Authorized User may then order, receive and use the requested materials, but must do so according to the statements and representations made in the application, and any conditions set forth by the safety committee and all applicable local, state and federal laws, regulations and license conditions. Violations or infractions of these conditions may be cause for suspension or termination of the approval to receive and use radioisotopes.

14.1.7. Approval will be given by the Committee for a period of three years.

14.2. Amendment

14.2.1. Any changes to the authorization, such as additional personnel, new radionuclide(s), increased or decreased possession limits, changes in experimental procedures which have an impact on safety, addition or removal of areas of radioactive material use, changes to the number of subjects or controls, or changes in the chemical or physical form of a material previously approved must be submitted as a request for amendment. [See the [Committee 1 Amendment to Authorization form](#) on the EHSO website.] Submit completed amendment forms to the Radiation Safety Officer.

RAD-030, RADIATION SAFETY MANUAL

- 14.2.2. Approval is granted by unanimous vote by a convened quorum after resolution of any comments or questions.
- 14.2.2.1. An administrative approval of changes of personnel to use isotopes, and addition or deletion of laboratory locations can be made at the discretion of the Radiation Safety Office.
- 14.2.2.2. Committee members receive a list of these administrative changes at each meeting for final approval.
- 14.2.3. Radiation Safety Committee 1 may require additional conditions under which the use of radioactive material must be conducted.
- 14.2.4. Prior to using radioactive material in a clinical area, the Radiation Safety Office will review the facility layout and proposed uses to determine if shielding is adequate to comply with exposure limits for the general public in adjacent areas.
- 14.2.5. An amendment must be submitted and approved prior to allowing any person to assume the responsibilities of an Authorized User, Authorized Medical Physicist, or Nuclear Medicine Technologist, as stipulated [in section 4.3](#).

14.3. Renewal

EHSO will contact the Authorized User to review the authorization prior to renewal. Any changes made during the review will be included in the renewal application. Renewal applications received by the expiration date of the authorization will allow the work to continue in the interim until Radiation Safety Committee 1 issues a new authorization.

14.4. Suspension

Radiation Safety Committee 1 and the Radiation Safety Office have the right to suspend an authorization, in full or in part, following the discovery of violations or infractions of the authorization. The Committee may reinstate the authorization once the violations have been corrected and an increased surveillance program put into place.

15. CLOSING/TERMINATING A LAB

The Radiation Permit Holder (RPH or Authorized User) may request assistance from the Radiation Safety Office when leaving a lab. However, it is the responsibility of the permit holder when leaving a laboratory to properly dispose of or safely remove all hazardous materials, waste and contaminated equipment in their lab. Work on all study protocols must cease. All laboratory spaces and equipment will be surveyed for radioactive contamination by EHSO. Guidelines can be found at [Laboratory Decommissioning Guidelines](#).

The authorization may be terminated upon the following conditions:

- 15.1.1. Written request of permit holder
- 15.1.2. Failure to renew authorization in a timely fashion
- 15.1.3. Violations or infractions of the authorization that are not corrected

RAD-030, RADIATION SAFETY MANUAL**16. FOR MACHINE-PRODUCED RADIATION IN RESEARCH****16.1. Human Use and Clinical Trials**

- 16.1.1. The Radiation Safety Committee 3 reviews applications for research that involve the use of ionizing radiation on the research subjects. New applications that are submitted to the Institutional Review Board (IRB) are automatically forwarded to the Radiation Safety Office. Alternatively, the application may be provided directly to the Radiation Safety Office.
- 16.1.2. The applicant should submit a copy of the Informed Consent, a Lay Summary, if available, and a [Radiation Summary Form](#), and any other documentation subsequently requested by the committee, to the Radiation Safety Office.
- 16.1.3. The Radiation Summary Form is a summary of radiological procedures that will be used in the research study. A copy of this form can be found on the website. The Radiation Permit Holder will indicate which procedures are research-driven and which are standard of care. The form also assists the RPH in estimating the total radiation dose likely to be given to the patient. This total dose is then categorized into Low Dose, Moderate Dose, or Significant Dose.
- 16.1.3.1. Standard of care includes imaging that is normally performed as indicated by the subject's medical provider for diagnosis or treatment of a disease. The sole criteria when making a decision concerning performing these scans is the benefit of the patient. (Note, these scans may have research benefit but this must not be a factor when considering performing the scan.) Any other scans are considered beyond Standard of Care.
- 16.1.4. The Radiation Safety Committee 3 will also review the Informed Consent to determine if the research subject is appropriately notified of the risks from the radiation. The applicant should include radiation risk information in the Consent that provides the subject the following three pieces of information - which procedures involve ionizing radiation; if the radiation is necessary for their care or if it is being done only for the research purposes; and a comparative estimate of the risk of cancer from these procedures. Guidance for such language can be found at [Consent Guidance](#).
- 16.1.5. Once all documents are submitted, the Radiation Safety Office will review them for completeness and forward them to the Radiation Safety Committee 3. The Radiation Safety Committee 3 is given sufficient time to review and comment.
- 16.1.6. RPH is contacted to resolve any comments or questions.
- 16.1.7. Approval is granted by unanimous vote by a convened quorum after resolution of any comments or questions.
- 16.1.7.1. Studies that involve only standard of care or low-dose procedures on adults are reviewed by the Committee 3 Chairman and the Radiation Safety Officer and registered as compliant with the requirements for informed consent rather than approved by committee, unless deferred to Committee 1. If either has questions the application may be referred to the full-committee at their discretion.
- 16.1.7.2. Studies that also involve only standard of care or low-dose nuclear medicine procedures on adults will be reviewed by the Committee 1 Chairman and the Radiation Safety Officer and registered as compliant with the requirements for informed consent rather than approved by committee. If either has questions the application may be referred to the full-committee at their discretion.

RAD-030, RADIATION SAFETY MANUAL

16.1.8. The approval document is either sent to the IRB or to the RPH or research coordinator.

16.2. Non-Human Use

Any research use of machine-produced radiation on animals or samples requires an application to the Radiation Safety Committee 2. A copy of this application can be found at the [X-Ray Application Non-Human Use form](#).

17. FOR UNESCORTED ACCESS TO IRRADIATORS**17.1. How Do I Get Access To An Irradiator?**

17.1.1. Any person requesting to have unescorted access privileges to an irradiator must contact the [Radiation Safety Office](#) for further instructions. Trustworthiness and reliability (T&R) must be verified before unescorted access privileges will be granted. The Emory T&R Official or designee will provide all forms and advise applicants via email of next steps.

17.1.2. The approval process consists of the following steps:

17.1.2.1. Certification by Radiation Permit Holder: The RPH certifies that the applicant requires unescorted access to an irradiator.

17.1.2.2. Payment Approval for Fingerprinting Charges: The RPH provides an active SmartKey code for charges. The cost per applicant is \$75.00 for fingerprinting, plus other reasonable charges for verifying educational history.

17.1.2.3. T&R Verification: RPH may review past work history for employees who have been at Emory longer than three years. Work and education history will be reviewed by the T&R Official for employees who have been at Emory less than three years. This may include review of graduate school applications, Human Resources documentation, and other resources as required. Past employers may be contacted.

17.1.2.4. Fingerprinting: Digital fingerprints are taken by Emory Police Department and transmitted to the Nuclear Regulatory Commission for a criminal background check.

17.1.3. The Emory T&R Official or designee will provide specific approval and instructions for access once all steps are complete. Other users with unescorted access privileges may provide instruction on the operation of the irradiator only.

17.1.4. Access privileges can be revoked at any time for security violations and may only be reinstated following review and agreement by the T&R Official and the user's RPH.

17.2. But I Need Access Now!

Users approved for unescorted access may escort other users by completing the "Escorted Irradiator Access Documentation" Form and returning it to EHSO in accordance with the instructions on the form. The form is available upon request from the Emory T&R Official.

RAD-030, RADIATION SAFETY MANUAL**SAFE USE OF RADIOACTIVE MATERIAL****18. GENERAL RULES FOR HANDLING RADIOACTIVE MATERIAL**

Do not allow children under 18 years of age in laboratories where radioactive materials are used or stored unless they are students or employees of the University who have been approved by Radiation Safety.

18.1. Personal Protective Equipment (PPE)

18.1.1. Wear a laboratory coat or other protective clothing, disposable gloves, close-toed shoes and eye protection at all times when using radioactive materials. [Also see [EHSO PPE Guideline](#)].

18.1.2. PPE such as lab coats and gloves should not be worn outside the laboratory.

18.2. Contamination Control

18.2.1. All work bench areas must be covered with absorbent paper. Absorbent paper must be checked for contamination after each use. Work with large volumes of radioactive material must be done on a tray.

18.2.2. Monitor hands, shoes, and clothing for contamination after each procedure or before leaving the area.

18.2.3. Do not eat, drink, smoke, apply cosmetics, or change contact lenses in any area where radioactive material is stored or used. [See also [EHSO policy on food and drink in laboratories](#)]

18.2.4. Do not store food, drink, or personal effects in areas where radioactive material is stored or used.

18.2.5. Never pipette by mouth.

18.2.6. Dispose of radioactive waste only in designated, Emory approved, labeled and properly shielded receptacles.

18.3. Exposure Control

18.3.1. Shielding materials must be available for specific isotopes used in the lab. Use lead shielding for gamma emitters; Plexiglas for high energy beta emitters.

18.3.2. Wear personnel monitoring devices, if required, at all times while in areas where radioactive materials are used or stored.

18.4. Additional Precautions in Shared or Multi-User Labs

In large community laboratories shared by radiation users and non-radiation users, care will have to be exercised to limit both contamination events and exposure concerns. Quantities of and procedures with radioactive materials used in the open laboratory will be limited according to possible hazards connected with its use.

18.4.1. Recommended possession limits for RPH permits in shared labs

18.4.1.1. 10 mCi limit for H-3, C-14, P-33, S-35, Cr-51 and Tc-99m

18.4.1.2. 1 mCi limit for Na-22, Na-24, P-32, Cl-36, and Ca-45

18.4.1.3. 100 μ Ci limit for I-125 and I-131

RAD-030, RADIATION SAFETY MANUAL

- 18.4.1.4. Approval may be granted for possession limits above those recommended on a case-by-case basis.
- 18.4.2. Radioactive materials must be adequately shielded on all sides to maintain exposures at less than 2 mR/hr at one foot from the source.
- 18.4.3. Use enclosed labs with the doors closed or a chemical fume hood for:
- 18.4.3.1. Amounts of stable, non-volatile radioactive materials which exceed the above recommended possession limits for RPH permits in shared labs
 - 18.4.3.2. Procedures which may result in the production of small amounts of radioactive aerosols such as micro centrifuges and speed-vacs. Request specific approval if your equipment must be used in shared spaces.
 - 18.4.3.3. All unstable and/or volatile radioactive materials must be used in chemical fume hoods; preferably within an enclosed lab to restrict traffic.
- 18.4.4. Labeling
- 18.4.4.1. All countertops where radioactive materials are used must be clearly defined and labeled with the radiation symbol.
 - 18.4.4.2. All hoods in which radioactive materials are used must be clearly labeled with the radiation symbol.
 - 18.4.4.3. All sinks in which radioactive material is introduced by disposal or by cleaning of contaminated lab ware must be clearly labeled with the radiation symbol.
 - 18.4.4.4. Centrifuges, incubators and speed-vacs and other multi-user equipment in which radioactive material is used must be clearly labeled with the radiation symbol.

19. LABORATORY REQUIREMENTS***19.1. Required Lab Features***

Basic construction of Emory laboratories is suitable for most combinations of radiotoxicity and possession limits requested. Laboratory use generally requires handling only low levels of non-volatile radioactive material. Laboratories where radioactive material is used or stored must possess the following features:

- 19.1.1. Smooth and nonabsorbent floors
- 19.1.2. Countertops impervious to forms of radioisotope being used.
- 19.1.3. Hand washing sink with impermeable surface and drain set level with or below sink level to allow for complete draining.
- 19.1.4. Negative air pressure in laboratories with one time pass-through ventilation.
- 19.1.5. Access to a GM Survey meter, unless only H-3 is used.
- 19.1.6. Access to a liquid scintillation counter and/or gamma counter.

19.2. Additional Features Dependent on Usage

- 19.2.1. Lab requirements may increase with higher possession limits or radiotoxicity of the material.

RAD-030, RADIATION SAFETY MANUAL

19.2.2. Additional facilities, such as chemical fume hoods, portable or built-in shielding may be necessary for isotopes with higher radiotoxicity, such as Ca-45, I-125 or I-131.

20. GENERAL RULES FOR STORAGE & SECURITY OF RADIOACTIVE MATERIAL

Secure all radioactive materials when it is not under the constant surveillance and immediate control of the user(s). Store all radioactive solutions in clearly labeled containers. Areas where radioactive materials are stored need specific safeguards to prevent unauthorized removal or access.

20.1. Self-Contained laboratory

Radiation Permit Holders who are in a self-contained laboratory area must choose a method of securing their stock solutions from the following options:

20.1.1. Secured in a locked refrigerator or freezer;

20.1.2. Keep lab locked when vacant, even for short periods of time;

20.1.3. Keep material in a lock box inside the refrigerator or freezer.

20.1.3.1. The RPH's choice will become an audit item. If material is found unsecured, use of a lock box will become mandatory.

20.2. Shared or Multi-User Labs

20.2.1. All stock material must be kept in a lock box inside refrigerator, freezer or cabinet. The box must be secured to prevent easy removal. Refrigerators and freezers must be locked at the end of the day

20.2.2. Radioactive samples may be stored in a locked refrigerator or freezer located in a hallway or alcove or in a locked box within. Refrigerators and freezers must be locked at the end of the day

21. CLINICAL AREAS

Since clinical areas typically use much higher activities of radioactive material, additional safeguards should be in place. Consideration should be given to the location of flood sources and radioactive patients with respect to well-counters, dose calibrators and gamma cameras. Waste containers and sharps containers may also need added shielding to minimize exposure to staff. A spill kit should also be available, and its whereabouts known to all persons handling radioactivity. (See also "[Use of Protective Equipment](#)" in the Clinical Use section of this manual).

22. RADIOACTIVE VOLATILES AND GASES

All unstable and/or volatile radioactive materials (e.g. H-3 borohydride or C-14 methyl iodide) must be used in enclosed laboratories equipped with exhaust fume hoods.

22.1. Special Requirements for Users of Radioiodine-125 and -131

22.1.1. Only Radiation Permit Holders and technicians with prior experience using large quantities (1 mCi or greater) of radioiodine will be authorized to perform iodinations.

22.1.2. Each lab in which an iodination is to be done will be cleared of all individuals except those actively participating in the procedure.

RAD-030, RADIATION SAFETY MANUAL

- 22.1.3. All work will be done in an approved hood with a minimum air flow of 100 linear feet per minute with a sash opening of one foot.
- 22.1.4. The hood will be exhausted to the outside - Laminar flow hoods recirculating air within the room will not be used.
- 22.1.5. Air samples will be taken in each lab during the use of 1 mCi or greater quantities of I-125 or I-131 until the research radiation safety liaison is satisfied that the vapors are contained during the procedure.
- 22.1.6. A mandatory radiation survey and wipe test for radioactive contamination is required after each use.
- 22.1.7. A baseline thyroid bioassay should be performed prior to an iodination. A thyroid bioassay is required within 24 to 72 hours following an iodination exceeding 1 mCi I-125 or I-131. These are scheduled by appointment with the research radiation safety liaison.

23. SIGNS & POSTINGS**23.1. *What Signs Do I Need To Post And Where?***

If you are in a laboratory at Emory University and need signs, EHSO will provide the required postings for all hazards. Refer to the [Signage Program](#) to request the signs. The following is a list of signs required for radioactive material use:

- 23.1.1. "Caution, Radioactive Material" sign must be posted on the door of any area where radioactive material is used or stored.
- 23.1.2. "Restricted Area, Authorized Entrance Only", and "No Eating, Drinking or Smoking" signs may also be present, if available, depending on the building signage system.
- 23.1.3. "Just In Time" brochure with radiation safety emergency phone numbers must be posted conspicuously in areas of use, so that it is readily available to workers in case of emergencies. "Notice to Employees" sign from the Georgia Department of Natural Resources must be posted in areas to permit employees working in or frequenting any portion of a controlled area to observe a copy on the way to or from the place of employment.
- 23.1.4. "Caution: Radiation Area" sign must be posted at any area where a person may receive 5 mrem in an hour.
- 23.1.5. Other areas with higher radiation exposure rates or airborne radioactivity might require additional or different signage. Consult the Radiation Safety Office if these situations apply.

23.2. *Posting Exceptions*

The following areas are exempted from posting requirements:

- 23.2.1. Areas where radioactive material is only present for less than 8 hours, provided that it is constantly attended by an individual who is taking measures to prevent exposure to other persons in excess of regulatory limits.
- 23.2.2. Hospital areas with radioactive patients; provided that the patients are authorized for release.
- 23.2.3. Areas where only a sealed source is stored, provided that the dose rate at 30 cm from the source (or source housing) is less than 5 mrem/hr.

RAD-030, RADIATION SAFETY MANUAL**23.3. Labeling Requirements**

Stock containers of Radioactive Material shall bear a label displaying the radiation symbol and the words, "Caution, Radioactive Material", radionuclide, quantity of radioactivity and date of assay. Labeling of instruments, trays or racks containing samples is acceptable. Waste containers are labeled the same except date is added when full.

Also, see additional labeling requirements for open laboratories in [Laboratory Requirements](#).

24. WIPE TESTS AND GEIGER SURVEYS

Both area surveys and wipe tests are performed to help keep worker exposure ALARA and to demonstrate compliance with regulatory limits to the general public.

24.1. Geiger Surveys

Geiger counters (GM) are portable instruments used to detect ionizing radiation and can also be used to survey areas for ambient radiation dose rates ("area surveys"), providing the correct detector is used.

The Geiger counter is the least expensive, fastest and generally the most reliable means of detecting and measuring radioactive contamination. The beta pancake detector is used with the Geiger counter for finding and measuring beta radiation, and will detect all beta radioisotopes used at Emory except H-3 and Ni-63. It does not detect those nuclides because their betas are too low in energy to penetrate the window of the detector. Radioisotopes which may be detected with the beta pancake are C-14, S-35, P-33, P-32, Ca-45, Cl-36 and other beta emitting nuclides.

The low energy gamma probe is used with the Geiger counter to detect and measure gamma radioisotopes of various energies. It is most efficient for I-125, but will perform adequately for Cr-51, In-111, Co-57 and other gamma emitting nuclides. These detectors will also detect low energy x-rays, such as those emitted by beta emitters producing Bremsstrahlung radiation.

24.2. Wipe Tests

Wipe tests are performed to detect and quantify radioactive contamination on surfaces of work areas and/or equipment. Removable contamination can be detected and measured through a wipe test of the surface, which is counted in an appropriate counting instrument, such as a liquid scintillation counter, a sodium iodide or germanium gamma counter, or a proportional alpha/beta counter.

24.3. When to Perform

Dose-rate surveys, at a minimum, must be performed in locations where workers are exposed to radiation levels that might result in radiation doses in excess of 2 mrem/hr. Contamination surveys must be performed:

24.3.1. To evaluate radioactive contamination that could be present on surfaces of floors, walls, laboratory furniture, and equipment ;

24.3.2. After any spill or contamination event;

24.3.3. When procedures or processes have changed;

24.3.4. To evaluate contamination of users and the immediate work area, at the end of the day, when radioactive material is used;

RAD-030, RADIATION SAFETY MANUAL

24.3.5. In areas adjacent to restricted areas and in all areas through which radioactive materials are transferred and temporarily stored before shipment.

24.4. How to Perform

24.4.1. To ensure achieving the required sensitivity of measurements, survey samples will be analyzed in a low-background area.

24.4.2. A gamma counter system with a single or multi-channel analyzer can be used to count samples containing gamma-emitters (e.g., cesium-137, cobalt-60).

24.4.3. A liquid scintillation or gas-flow proportional counting system can be used to count samples containing beta-emitters and gamma-emitters. The instrument must be sufficiently sensitive to detect the presence of 200 dpm/100 cm² of removable contamination. Results must be documented in disintegrations per minute (dpm) on a facility diagram.

24.4.4. A radioactive source with a known amount of activity should be used to convert sample measurements (usually in counts per minute (cpm)) to dpm.

24.4.5. Each laboratory using gamma-emitting or high energy beta-emitting material must have a suitable survey meter available. The survey instrument must be checked for consistent response with a dedicated source before each use. Do not use an instrument that does not respond appropriately to the source. It is not necessary to keep records of these checks.

24.5. Required Frequency

When working with radioactive material, hands, lab coats (especially edge of sleeves), and shoes should be monitored before leaving the laboratory both during the work day and at the end of the work day before leaving the workplace.

Perform radiation level surveys with a survey meter sufficiently sensitive to detect 0.1 milliroentgen per hour (mR/hr) in the following areas, at the frequency specified:

24.5.1. Survey at the end of each day of use all radiopharmaceutical elution, preparation, assay and administration areas (except patient rooms, which will be surveyed at the end of the therapy instead of on the day of administration) when using radiopharmaceuticals requiring a written directive (e.g., all therapy dosages and any iodine-131 dosage exceeding 30 μ Ci).

24.5.2. Survey weekly all radionuclide use, storage, and waste storage areas. If diagnostic administrations are occasionally made in patients' rooms (e.g., bone scan injections, Tc-99m heart agents) and special care is taken to remove all paraphernalia, those rooms need not be surveyed.

24.5.3. Survey during the semi-annual inventory all sealed-source and brachytherapy-source storage areas.

24.5.4. Removable contamination survey samples should be measured in a low-background area. The following areas and frequencies should be followed:

24.5.5. Perform removable contamination surveys weekly for:

24.5.5.1. laboratory areas where only small quantities radioactive material are used (<1 millicurie at a time), or

24.5.5.2. radionuclide storage and radionuclide waste storage areas, or

RAD-030, RADIATION SAFETY MANUAL

24.5.5.3. Areas where radiopharmaceuticals with half-lives of less than 2 hours are eluted, prepared, assayed, or administered.

24.5.6. Perform removable contamination surveys daily for areas where generators and multi-use bulk vials of radiopharmaceuticals with half-lives of more than 2 hours are eluted, prepared, or assayed.

24.5.7. Perform removable contamination surveys weekly for areas where radiopharmaceuticals are administered.

24.5.8. If diagnostic administrations are occasionally made in patients' rooms (e.g., bone scan injections, Tc-99m heart agents), with special care taken to remove all paraphernalia, those rooms need not be surveyed.

24.6. Action Levels – What Are They & What If They Are Exceeded

The tables below list the action levels for area surveys and wipe tests.

Table 4, Action Level for Ambient Surveys

ACTION LEVEL FOR AMBIENT SURVEYS:	
Unrestricted areas	> 2 mR/hr @ 30 cm > 2 mR/hr @ contact for personal clothing & skin
Restricted areas	> 10 mR/hour @ 30 cm (protective clothing used only in restricted area)

Table 5, Wipe Test Action Levels - Unrestricted Areas

WIPE TEST ACTION LEVELS IN UNRESTRICTED AREAS (dpm/100 cm ²)	
Nuclide	Removable
Any beta or gamma emitters	200 dpm/100 cm ²

Table 6, Wipe Test Action Levels - Restricted Areas

WIPE TEST ACTION LEVELS IN RESTRICTED AREAS (dpm/100 cm ²)	
Nuclide	Removable
Any beta or gamma emitters	2000 dpm/100 cm ²

RAD-030, RADIATION SAFETY MANUAL

- 24.6.1. If action levels for dose rates are exceeded, you must shield the source in order to reduce the dose rate to within the action level. You must notify the Radiation Safety Office if you discover an unrestricted area has received more than 2 mrem in an hour.
- 24.6.2. If action levels for removable contamination are exceeded, then you must perform decontamination steps until the amount of contamination is below the action level. Contamination found in unrestricted areas and on personal clothing will be immediately decontaminated to background levels. The worker must retain a copy of the survey in the laboratory radiation records documenting that an effort was made to reduce ambient radiation levels through decontamination or shielding.
- 24.6.3. If the dose rates cannot be brought below the action levels:
- 24.6.3.1. Shield the radiation
 - 24.6.3.2. Post the area to restrict its access.
 - 24.6.3.3. Notify the Radiation Safety Office.
 - 24.6.3.4. Assist in investigating what caused the action level to be exceeded. Records of dose-rate surveys and wipe tests must include the following:
 - 24.6.3.5. A diagram of the area surveyed or a list of items and equipment surveyed
 - 24.6.3.6. Specific locations on the survey diagram where wipe test was taken;
 - 24.6.3.7. Measured dose rates in mR/hr and contamination levels in dpm/100 cm²;
 - 24.6.3.8. Name or initials of the person who conducted the survey and the date;
 - 24.6.3.9. Make and model number of equipment used;
 - 24.6.3.10. Action taken in case of excessive dose rate or contamination, and follow-up survey information.

25. SEALED SOURCES

Sealed Sources or source holders must identify isotope, activity and date of assay of the source. Sources must be stored in such a manner that the dose rate in adjacent unrestricted areas does not exceed 2 mrem in one hour and 50 mrem in one year.

25.1. Leak-Tests & Inventory of Sealed Sources

Any source of 100 μCi or more (or 10 μCi or more for sources designed to emit alpha particles) with a half-life greater than 30 days (excluding H-3) must be tested for leakage or contamination prior to initial use, any time there is reason to suspect that the source might be leaking, and at least every six months. Sealed sources will be leak-tested by Radiation Safety personnel. Leak tests are not required on sources that are stored and are not being used, but are required before use or transfer.

All sealed sources will be inventoried by Radiation Safety every six months.

RAD-030, RADIATION SAFETY MANUAL

Table 7, Radionuclides Commonly Used in Research Laboratories at Emory University

ISOTOPE	BETA/ GAMMA	PERSONNEL MONITORING	SHIELDING	GM METER	HAZARD	LABORATORY
H-3	Beta	None	None	No	Low	Up to 10 Ci
C-14	Beta	None	None	Yes	Moderate	Up to 100 mCi
F-18	Positron/ Gamma	Yes	Lead	Yes	Moderate	Up to 100 mCi
Na-22	Gamma	Yes	Lead	Yes	High	Up to 1 mCi, med-100 mCi
Na-24	Gamma	Yes	Lead	Yes	Moderate	Up to 100 mCi
P-32	Beta	wb>50 mCi ring >1mCi	Plexiglas	Yes	Moderate	Up to 100 mCi
P-33	Beta	None	None	Yes	Moderate	Up to 100 mCi
S-35	Beta	None	None	Yes	Moderate	Up to 100 mCi
Cl-36	Beta	None	None	Yes	High	Up to 1 mCi, med-100 mCi
Ca-45	Beta	None	None	Yes	High	Up to 1 mCi, med-100 mCi
Mn-54	Gamma	Yes	Lead	Yes	Moderate	Up to 10 mCi
Fe-55	Electron Capture	None	None	No	Low	Up to 10mCi
Rb-86	Beta	Yes	Plexiglas	Yes	Moderate	Up to 100mCi
Cu-64	Gamma	Yes	Lead	Yes	Moderate	Up to 100mCi
Y-90	Beta	Yes	Plexiglas	Yes	Moderate	Up to 100 mCi
Cr-51	Gamma	wb>10 mCi	Lead	Yes	Moderate	Up to 100 mCi
Tc-99m	Gamma	Yes	Lead	Yes	Low	Up to 10 Ci
In-111	Gamma	Yes	Lead	Yes	Moderate	Up to 100 mCi
I-123 I-125 I-131	Gamma	> 1 mCi	Lead	Yes	High	Up to 1 mCi, med-100 mCi

H-3, S-35 and the radioiodines have more volatile forms.

Use of a chemical fume hood and bioassay may be required on a case-by-case basis

RAD-030, RADIATION SAFETY MANUAL**ACCIDENTS / INCIDENTS / EMERGENCIES**

Emory University's [Just In Time Guide To Campus Emergencies](#) is a quick reference tool that provides helpful and quick information for students, faculty and staff to use in the event of an emergency. Excerpts of the guide that pertain to the use of radioactive material are listed below for convenience.

Emory Healthcare (EHC) personnel receive training in Emergency Preparedness and the use of the EHC Unified Emergency Codes. However, they should review the procedures below that are appropriate to their use of radioactive material.

26. PERSONAL CONTAMINATION**26.1. What If I Am Contaminated?**

- 26.1.1. Any personal contamination must be reported to the Radiation Safety Office immediately.
- 26.1.2. Contaminated skin should be washed with mild soap and water.
- 26.1.3. Contaminated clothing must be removed promptly and folded inward to prevent the spread of contamination. The clothing should then be placed in a plastic bag and labeled for decay or disposal as radioactive waste by radiation safety.
- 26.1.4. The Radiation Safety Office will record contamination levels observed and procedures followed for incidents involving contamination of individuals. An incident record will be documented that includes names of individuals involved, description of work activities, calculated dose, probable causes (including root causes), steps taken to reduce future incidents of contamination, times and dates, and the surveyor's signature.

26.2. What If I Am Injured?

IN THE EVENT PERSONNEL ARE INJURED, SEEK MEDICAL ATTENTION IMMEDIATELY!

Employee incidents are reported to Employee Health. Inform your supervisor, staff leadership, or safety representative.

26.3. How Do I Reach Medical Help?

- 26.3.1. If this is an Emergency, call 911
 - 26.3.1.1. Page the on-call Employee Health/Occupational Injury Management Nurse Practitioner for further guidance at: 404-686-5500, PIC# 50464.
 - 26.3.1.2. Notify the employee's supervisor or their designee within 24 hours of the incident.
- 26.3.2. Emory Healthcare
 - 26.3.2.1. Access the Emory Healthcare Intranet page and go to the Occupational Injury Management webpage <http://www.ouehc.org/departments/human-resources/occupational-injury-management>.
 - 26.3.2.2. Click on protocols, emergency treatment and medication management.

RAD-030, RADIATION SAFETY MANUAL

26.3.2.3. Click on protocol-injury and follow these instructions.

<http://www.ourehc.org/departments/human-resources/occupational-injury-management/employee-incident-protocol.pdf>.

26.3.3. EXPRESS CARE CLINIC

26.3.3.1. The Express Care Clinic provides all Emory Healthcare and Emory University employees care for acute occupational accidents and exposures.

26.3.3.2. Accident/Injury Reports must be completed within 24 hours of the incident. See next section.

26.4. *How to Record Incident and Exposures?*

26.4.1. Emory Healthcare patient and visitor incidents and near misses are reported in the SAFE Incident Reporting System. Access the Emory Healthcare Intranet page, <http://www.ourehc.org>. Click on SAFE Incident Reporting under quick links to complete a Patient/Visitor SAFE Event Report.

26.4.2. Emory Healthcare employees can report incidents in e-Vantage. Access the Emory Healthcare Intranet page, <http://www.ourehc.org>. Click on SAFE Incident Reporting under quick links and follow the instructions under Employee Injury/Illness Report.

26.4.3. University event reporting system can be accessed through the 'How to Report an Accident' section on the [Environmental, Health, and Safety Office \(EHSO\) website](#). If the employee is unable to complete the report it should be completed by the supervisor.

27. WHAT IF I HAVE A SPILL?

The decision to implement a major spill procedure instead of a minor spill procedure depends on many incident specific variables, such as the number of individuals affected; other hazards present; the likelihood of spread of contamination; and types of surfaces contaminated as well as the radiotoxicity of the spilled material. For some spills of short-lived radionuclides, the best spill procedure may be restricted access pending complete decay.

As a general guideline, a spill involving more than one millicurie of radioactive material or more than one liter of radioactive liquid is a major spill which must be reported immediately to the Radiation Safety Office. The initial responder can determine if the clean-up will require additional radiation safety assistance.

27.1. *Minor Spills of Liquids and Solids*

27.1.1. Notify persons in the area that a spill has occurred.

27.1.2. Prevent the spread of contamination by covering the spill with absorbent paper. (Paper should be dampened if solids are spilled.)

27.1.3. Clean up the spill, wearing disposable gloves and using absorbent paper.

27.1.4. Carefully fold the absorbent paper with the clean side out and place in a plastic bag for transfer to a radioactive waste container. Put contaminated gloves and any other contaminated disposable material in the bag.

27.1.5. Survey the area with an appropriate radiation survey meter set on lower scale. Check the area around the spill for contamination. Also check hands, clothing, and shoes for contamination.

27.1.6. Report the incident to the Radiation Safety Office promptly.

RAD-030, RADIATION SAFETY MANUAL

27.1.7. Cooperate with radiation safety personnel in discovering the root cause of the spill and in providing requested bioassay samples if indicated.

27.2. Major Spills of Liquids and Solids

27.2.1. Notify all persons not involved in the spill to clear the room, but to remain in the area to await survey.

27.2.2. Prevent the spread of contamination by covering the spill with absorbent paper (paper should be dampened if solids are spilled), but do not attempt to clean it up. To prevent the spread of contamination, limit the movement of all personnel who may be contaminated.

27.2.3. Shield the source only if it can be done without further contamination or significant increase in radiation exposure.

27.2.4. Close the room and lock or otherwise secure the area to prevent entry. Post the room with a sign to warn anyone trying to enter that a spill of radioactive material has occurred.

27.2.5. Notify the Radiation Safety Office immediately.

27.2.6. Survey all personnel who could possibly have been contaminated. Decontaminate personnel by removing contaminated clothing and flushing contaminated skin with lukewarm water and then washing with a mild soap.

27.2.7. Allow no one to return to work in the area unless approved by radiation safety personnel.

27.2.8. Cooperate with radiation safety personnel in discovering the root cause of the spill and in providing requested bioassay samples if indicated.

27.2.9. Follow the instructions of the Radiation Safety Office staff concerning decontamination techniques, surveys, provision of bioassay samples and requested documentation.

28. OTHER INCIDENTS**28.1. Incidents Involving Dust, Mist, Fumes, Organic Vapors or Gases**

28.1.1. Notify all personnel to vacate the room immediately.

28.1.2. Shut down ventilation system, if possible, unless it is determined that the room ventilation system needs to be used to clear the air for access purposes.

28.1.3. Vacate the room. Seal the area, if possible.

28.1.4. Notify the Radiation Safety Office immediately.

28.1.5. Ensure that all access doors to the area are closed and posted with radiation warning signs or post guards (trained) at all access doors to prevent accidental opening of the doors or entry to the area.

28.1.6. Survey all persons who could have possibly been contaminated. Decontaminate as directed by Radiation Safety Office personnel.

28.1.7. Promptly report suspected inhalation and ingestions of radioactive material to the Radiation Safety Office.

28.1.8. Decontaminate the area only when advised and/or supervised by the Radiation Safety Office.

28.1.9. Allow no one to return to work in the area unless approved by the Radiation Safety Office.

RAD-030, RADIATION SAFETY MANUAL

- 28.1.10. Cooperate with radiation safety personnel in discovering the root cause of the incident and in providing requested bioassay samples if indicated.
- 28.1.11. Follow the instructions of the Radiation Safety Office staff concerning decontamination techniques, provision and collection of bioassay samples, and providing requested documentation.

28.2. *Minor Fires*

- 28.2.1. Immediately attempt to put out the fire by approved methods (e.g., fire extinguisher) if other fire hazards or radiation hazards are not present.
- 28.2.2. Notify all persons present to vacate the area and have one individual immediately call the fire department and Radiation Safety Office.
- 28.2.3. Once the fire is out, isolate the area to prevent the spread of possible contamination.
- 28.2.4. Survey all persons involved in combating the fire for possible contamination.
- 28.2.5. Decontaminate personnel by removing contaminated clothing and flushing contaminated skin with lukewarm water, then washing with a mild soap.
- 28.2.6. In consultation with Radiation Safety Office, determine a plan of decontamination and the types of protective devices and survey equipment that will be necessary to decontaminate the area.
- 28.2.7. Allow no one to return to work in the area unless approved by the Radiation Safety Office.
- 28.2.8. Cooperate with radiation safety personnel in discovering the root cause of the incident and in providing requested bioassay samples if indicated.
- 28.2.9. Follow the instructions of the Radiation Safety Office staff concerning decontamination techniques, provision and collection of bioassay samples, and providing requested documentation.

28.3. *Major Fire, Explosion or Major Emergencies*

- 28.3.1. Notify all persons to vacate the area immediately.
- 28.3.2. Notify the fire department.
- 28.3.3. Notify the Radiation Safety Office and other facility safety personnel.
- 28.3.4. Upon arrival of firefighters, inform them where radioactive materials are stored or where radioisotopes were being used; inform them of the present location of the radioactive material and the best possible entrance route to the radiation area, as well as any precautions to avoid exposure or risk of creating radioactive contamination by use of high pressure water, etc.
- 28.3.5. Cooperate with radiation safety personnel in discovering the root cause of the incident and in providing requested bioassay samples if indicted.
- 28.3.6. Allow no one to return to work in the area unless approved by the Radiation Safety Office.
- 28.3.7. Follow the instructions of the Radiation Safety Office staff concerning decontamination techniques, provision and collection of bioassay samples, and providing requested documentation.

RAD-030, RADIATION SAFETY MANUAL**28.4. *Release into Environment***

Immediately report to the Radiation Safety Office any unplanned release of radioactive material into the environment. The Radiation Safety Office will determine, based on the quantity released, if the event is required to be reported to the Georgia DNR.

28.5. *What If Radioactive Material Is Missing?*

Once a loss of radioactive material has been discovered, it must be immediately reported to the Radiation Safety Office. The Radiation Safety Office will:

28.5.1. Gather information regarding the disappearance of the radioactive material;

28.5.2. Initiate steps to locate and recover the material;

28.5.3. Determine if the loss is required to be reported to the Georgia DNR according to regulations, and

28.5.4. If required, report the loss in the required time frame.

RAD-030, RADIATION SAFETY MANUAL**PROCUREMENT****29. PROCUREMENT OF RADIOACTIVITY**

Radioactive materials (except for clinical uses) must be procured from NRC or Agreement State licensed suppliers via the Emory Express online system. Each supplier of radioactive material is required by state and federal regulations to possess a valid license to manufacture and prepare such material. The Radiation Safety Office will provide copies of the broad scope license to suppliers. If a copy is needed by a new or prospective supplier, the Radiation Permit Holder must advise the Radiation Safety Office for a copy to be forwarded to the supplier.

29.1. *How to Order Radioactive Material for Research Laboratories*

29.1.1. All Requisitions are submitted using Emory Express online ordering system. A Radiation Permit Holder can request radioactive material from a hosted catalog or as a non-catalog item. Please visit <http://www.finance.emory.edu> and follow guidelines for ordering radioactive material. All orders must be approved by EHSO prior to being processed by the Purchasing Department.

29.1.2. Orders cannot be placed directly with the vendor.

29.1.3. The Radiation Permit Holder must be authorized to use the isotopes and amounts prior to submitting an order.

29.1.4. Replacements for an incorrect order or unusable shipments must be negotiated by the Purchasing department with consent of EHSO.

29.1.5. The requisition must contain the name of the Radiation Permit Holder, the authorization number and the following product information:

29.1.5.1. Vendor

29.1.5.2. Catalogue number

29.1.5.3. Isotope

29.1.5.4. Compound

29.1.5.5. Activity in μCi or mCi

29.1.5.6. Quantity (number of units)

29.1.6. After verification by EHSO personnel that the Radiation Permit Holder is authorized to possess the material and that the order will not exceed the possession limit, the order will be transmitted to Purchasing by EHSO.

29.2. *How to Transfer from Other Universities/ Vendor Replacements*

If possible, a requisition with zero cost entered should be submitted via Emory Express. If this is not possible, a [Non-Emory Express Acquisition Form](#) must be completed by the receiving Radiation Permit Holder and submitted to EHSO for signature approval before the material is shipped.

29.3. *How to Transfer Within the University to Another Lab*

29.3.1. An [Emory University Radioactive Material Transfer Form](#), completed and signed by both parties involved in the transfer, must be submitted to EHSO for signature approval before the material is transferred.

RAD-030, RADIATION SAFETY MANUAL

29.3.2. Material must not be released until the transferor receives the transfer form approved by Radiation Safety personnel.

29.3.3. The transfer must be approved by Radiation Safety Office prior to transfer of the Material.

29.3.4. Transport of Radioactive material must be in compliance with all DOT/IATA regulations. Contact Radiation Safety Office for information on assistance in hazardous materials transportation.

29.4. *How to Order Sealed Sources*

29.4.1. Sealed sources for clinical areas are shipped directly to the departments, except for patient therapy sources for Radiation Oncology.

29.4.2. All other sealed sources must be received and processed by Radiation Safety Office.

29.5. *Ordering Radioactive Material for Clinical Departments*

29.5.1. Routine orders for radiopharmaceuticals are placed by the Nuclear Medicine Technologist or Medical Physicist. Nuclear pharmacy personnel will deliver the radiopharmaceutical directly to the nuclear medicine hot lab which they are able to access and secure.

29.5.2. When ordering radioactive materials for therapeutic uses, the person ordering the material will reference the Authorized User's written request when placing the order. The Authorized User's request will indicate the isotope, compound or physical form and activity level.

RAD-030, RADIATION SAFETY MANUAL**RECEIVING RADIOACTIVE MATERIAL****30. RECEIVING PACKAGES OF RADIOACTIVITY**

EHSO will receive packages for laboratories and check them for leakage and proper contents. EHSO will deliver the packages to the labs and will ask for a signature of receipt. For delivery of short-lived materials directly to the laboratory see [below](#). All material procurement must be as described in [Procurement](#).

30.1. Procedure for Receiving Packages

All personnel involved with the receipt of radioactive material shipments must be instructed in the proper procedures and precautions. Unpacking of Radioactive Materials must be done in accordance with established policy and procedures. See [Procedure for Receiving and Opening Packages](#).

- 30.1.1. Appropriate information for packages processed by EHSO will be entered in the EHSAssistant database.
- 30.1.2. When material is delivered to a laboratory by Radiation Safety, a member of the laboratory must be present to sign for it. If no one is present in the preferred lab or another lab belonging to the same Radiation Permit Holder, the package is returned to Radiation Safety. An attempt will be made to deliver the package later in the day, or a voice-mail message will be left for the Radiation Permit Holder or laboratory personnel.
- 30.1.3. Labels (e.g., white I, yellow II) on shipping boxes used for receiving radioactive materials must be defaced prior to disposal through housekeeping.

30.2. Who Can Receive Radioactive Material Directly Without Going Through Radiation Safety?

- 30.2.1. Most laboratory authorizations are not approved to check-in packages containing radioactive material and must have all radioactive material shipments delivered to the Radiation Safety Office for check-in and processing.
- 30.2.2. Researchers receiving time-sensitive material for animal use from a local radiopharmacy may receive it directly by special permission and instruction from Radiation Safety personnel. Laboratories must document the package receipt inspection and must maintain the documentation for inspection.
- 30.2.3. Shipments of Radioactive Materials for clinical human use may be shipped directly to clinical departments and received during working hours.
- 30.2.4. Due to the short half-life, isotopes produced in the Emory or PETNet cyclotron are transferred directly to Radiation Permit Holders without being routed to Radiation Safety Office.
- 30.2.5. The transfer of labeled compounds between laboratories can be made within the limits specified in the Authorization but must be approved by Radiation Safety and documented. (See "[How To Transfer Radioactive Material within the University to Another Lab](#)").

RAD-030, RADIATION SAFETY MANUAL**MAINTENANCE OF RADIOACTIVE MATERIAL INVENTORY****31. RADIOACTIVE MATERIAL INVENTORY**

Upon receipt of Radioactive Materials, the Authorized User/RPH is responsible for maintaining accurate inventory records for all Radioactive Materials possessed. Materials received in the lab must be recorded in online Inventory System. All laboratory radioactive material users must have training on how to use the EHSAssistant database. Contact Radiation Safety Office to schedule training. Clinical areas are responsible for maintaining their own inventory.

31.1. What Radioactive Material Am I Supposed to Track?

31.1.1. Each lab must keep track of any use of radioactive material. Use is tracked in the EHSAssistant database, which the lab accesses on-line. Specifically:

31.1.1.1. the date radioactive material is removed from a stock vial,

31.1.1.2. the amount of radioactivity used (in milliliters), and

31.1.1.3. The waste stream (dry, liquid, or mixed with liquid scintillation fluid) that it is destined for are the items to be recorded.

31.1.2. When an order has been received by the EHSO staff, the radioactive stock vial information is added to the labs inventory in the EHSAssistant database. The activity and the volume, among other items, are recorded. EHSAssistant thus enables the labs to track their use of radioactivity by date and volume, eliminating the need to calculate radioactive decay.

31.2. How Do I Learn To Use EHSAssistant?

31.2.1. Database training is taught during Radiation Safety Training for Lab Workers;

31.2.2. The lab worker can also access a written tutorial at www.ehso.emory.edu.

31.2.3. The lab can always contact the Radiation Safety Office for assistance.

31.3. What If I Made a Mistake?

31.3.1. Anytime you are unable to correct a mistake yourself, please contact your Radiation Safety Office professional for assistance.

RAD-030, RADIATION SAFETY MANUAL**TRANSPORTING RADIOACTIVE MATERIAL****32. TRANSPORTATION OF RADIOACTIVITY FROM EMORY**

All radioactive material shipped from Emory University must be shipped by or under the direction of Radiation Safety. Clinical departments are generally authorized to return radioactive material to their vendors. However, only staff currently trained in DOT radioactive material regulations can package shipments for transport. As of the writing of this Manual, DOT training is required every 3 years for road transport, and IATA training is required every 2 years for shipping by air.

32.1. *Can I Drive Radioactive Material Around Myself?*

No.

32.2. *How Do I Ship Radioactive Material to Somewhere Else?*

- 32.2.1. Laboratories or other areas not specifically authorized for packaging and shipping of radioactive materials must contact Radiation Safety for assistance. Items must be brought to Radiation Safety in unsealed shipping cartons. Pick-up can be arranged for shipments from off-campus locations.
- 32.2.2. Before any radioactive material can be shipped to any licensed recipient, Emory University must have a copy the recipient's NRC or Agreement State radioactive materials license to ensure that they are authorized to receive the material in the quantity being sent. For shipments other than routine returns to vendors, a Transfer Form must also be completed which has been signed by the Radiation Safety Officer of the recipient indicating approval of the transfer.
- 32.2.3. Records of radioactive material shipments must be kept according to [Radiation Safety Office Record Keeping Guidelines](#).

RAD-030, RADIATION SAFETY MANUAL**WASTE MANAGEMENT****33. LABORATORY AREAS****33.1. *How Do I Collect Radioactive Waste Safely?***

33.1.1. Waste containers are provided by EHSO for use in the laboratory:

33.1.1.1. 20 gallon stapacks for dry waste,

33.1.1.2. 5 gallon plastic buckets for liquid scintillation vials or dry waste, and

33.1.1.3. 1 gallon plastic bottles for liquids.

33.1.1.4. Alternate containers for radioactive waste must be approved in advance by EHSO.

33.1.2. Label containers with:

33.1.2.1. the isotope(s) to be placed in the container,

33.1.2.2. the Radiation Permit Holder's name, and

33.1.2.3. container number (from EHSAssistant) prior to putting the container into use;

33.1.3. Remove or deface any radioactive markings from primary containers before placing material in waste container.

33.1.4. When sealed for transfer to EHSO, add to the label:

33.1.4.1. the amount of activity as calculated by EHSAssistant,

33.1.4.2. radiation exposure rate (in mR/hr) at surface of the container,

33.1.4.3. exposure rate (in mR/hr) at 1 meter from container,

33.1.4.4. name of surveyor, and

33.1.4.5. date of survey

33.1.5. Lead shielding must be disposed of with EHSO separate from all other waste and must never be placed in a radioactive waste container.

33.2. *How Should I Separate My Radioactive Waste?*

33.2.1. Mixed chemical and radioactive waste generation must be approved by EHSO prior to beginning the experiment.

33.2.2. Short-Lived Dry Waste (half-life < 120 days) must be placed in separate containers according to half-life. Examples of isotopes that must be in separate containers include P-32, S-35 and I-125.

33.2.3. Short-Lived Aqueous Waste (half-life <120 days) is collected in plastic bottles supplied by EHSO. Examples of isotopes that must be in separate containers include P-32, S-35 and I-125. Aqueous waste used with potentially biohazardous materials or capable of supporting bacterial growth must be disinfected by the lab before disposal.

33.2.4. Long-Lived Dry Waste (half-life >120 days) must be placed in a 20 gallon stapacks or 5 gallon pail. The primary components of this waste are H-3 and C-14.

33.2.5. Long-Lived Aqueous Waste (half-life > 120 days) is collected in plastic bottles provided by EHSO. The primary components of this waste are H-3 and C-14.

RAD-030, RADIATION SAFETY MANUAL**33.2.6. Liquid Scintillation Vials**

33.2.6.1. Vials are only collected in five gallon pails due to the weight.

33.2.6.2. Separate pails must be used for each isotope being collected.

33.2.7. Sharps must be stored in biohazard sharps containers.

33.2.8. Animal Waste

33.2.8.1. Cage/pen waste (i.e., other excreta and bedding) must be collected in appropriate biohazard bags, sealed with tape, and labeled with:

33.2.8.1.1. the contents,

33.2.8.1.2. Radiation Permit Holder's name,

33.2.8.1.3. nuclide,

33.2.8.1.4. total collected activity, and

33.2.8.1.5. date of collection

33.2.8.2. Bags must be placed in designated storage area.

33.2.8.3. Carcasses and tissues must be placed in appropriate biohazard bags (with double bagging for heavy carcasses), sealed with tape and labeled with:

33.2.8.3.1. contents,

33.2.8.3.2. the Radiation Permit Holder's name,

33.2.8.3.3. nuclide,

33.2.8.3.4. total activity, and

33.2.8.3.5. Date administered.

33.2.8.4. When double bagging is used, all labels must be on the outside of the outermost bag.

33.2.8.5. Carcasses and animal tissues must be frozen and stored in designated freezers until disposal can be arranged through EHSO.

33.3. *Can I Pour Liquid Waste Down The Drain?*

In-laboratory disposal via sanitary sewer is allowed with Radiation Safety Office approval only on a case-by-case basis, usually when research generates a large volume of low activity waste. Clinical areas are authorized for in-house waste disposal of short-lived radioactive waste. The protocol below must be followed.

33.3.1. Survey the container with a GM survey meter.

33.3.2. If the reading is background, the material is ready for disposal.

33.3.3. Otherwise, it is held until it reaches background.

33.3.4. Run water in the designated sink for several minutes to ensure that it is draining properly.

33.3.5. Use proper personal protective equipment (PPE) to prevent contamination including lab coat, double latex or nitrile gloves, disposable apron, disposable protective sleeves and face shield.

33.3.6. Discharge liquid waste slowly to minimize splashing with water running, to be sure that the material moves out of the sink and into the sewer system.

RAD-030, RADIATION SAFETY MANUAL

33.3.7. Triple-rinse each container with water.

33.3.8. Survey the sink and surrounding work surfaces to confirm that no residual material or slight contamination is present and decontaminate if necessary.

33.4. *How Do I Tell Radiation Safety My Waste Is Ready For Pick-Up?*

Researchers must submit a request in EHSAssistant to have waste picked up.

33.5. *When Will My Waste Be Picked-Up?*

Waste will be picked up on the daily schedule established for research buildings. This schedule can be found in EHSAssistant when submitting a waste pickup request.

33.6. *Can I Let My Waste Decay Away In My Lab?*

Storing waste in the lab for decay is discouraged. Radiation Safety will pick up, manage and document the disposal.

34. CLINICAL AREAS

Clinical areas and PET Imaging areas using short-lived radioactive material are generally authorized to store, hold, decay, and dispose of their own waste. Waste disposal must be performed according to [Clinical Radioactive Waste Disposal Procedure](#).

Clinical areas are encouraged to contact the Radiation Safety Office if they need assistance with large quantities of radioactive waste or waste with half-lives longer than ten days.

RAD-030, RADIATION SAFETY MANUAL**USE OF RADIOACTIVE MATERIAL IN ANIMALS****35. SAFE USE OF RADIOACTIVITY IN VIVO****35.1. General Guidance**

- 35.1.1. An “Amendment to Non-Human Use Authorization-Committee II” form, “Radioactive Material in vivo” form, and an IACUC protocol must be submitted to Radiation Safety Office for approval by Committee II. IACUC has the prerogative to approve the use of animals.
- 35.1.2. The Radiation Safety Office at Emory University performs quarterly radiation safety audits of all Animal Use areas. Records such as Animal Care Record Cards, Cage Cards, Survey Records, Waste Logs, and Training Records are checked during the Audit. The Radiation Safety staff also performs contamination and radiation level surveys if needed.
- 35.1.3. All use of radioactive material must be conducted under the supervision of a Radiation Permit Holder who has approval from Division of Animal Resources (DAR), Institutional Animal Care and Use Committee (IACUC), and Radiation Safety.
- 35.1.4. All investigators using radioisotopes in animals are required to design and perform their studies in a manner which prevents unnecessary exposures to radiation and keeps necessary exposures ALARA (As Low as Reasonably Achievable). This requirement applies to the use of all Animal Care Facilities.

35.2. Standard Procedures

- 35.2.1. All Animal Care Personnel and Research Staff must have current Radiation Safety Training on the safe use and sources of radiation. Personnel shall be aware of the significance of radioactive signs and labels and follow precautionary measures included on such signs.
- 35.2.2. Main entry doors of Animal Care Facilities where Radioactive Material is present must be posted with a sign bearing the radiation hazard symbol and the words “CAUTION RADIOACTIVE MATERIAL”.
- 35.2.3. Primary enclosures containing live animals that have received Radioactive Material must be properly identified in the Animal Care Facility. Cage Cards must be attached to the cage in which Radioactive Animals are housed. See Section 35.3 below for cage card selection.
- 35.2.4. Radioactive animals must be transported in such a manner to prevent contamination of hallways, elevators, etc. Solid bottomed transfer containers are mandatory.
- 35.2.5. The Radiation Permit Holder must evaluate the radiation dose in the work place, the excretion rate of radioactive material, and any special hazard that may be associated with the radionuclide or its chemical form.
- 35.2.6. If routine animal care is handled by staff other than the Radiation Permit Holder’s Staff, specific instructions must be provided on animal care such as feeding, watering, cage and pen cleaning, and waste handling.
- 35.2.7. In the event animals undergo necropsy during the period of radioactivity, procedures for radiation protection, sample collection, and waste disposal must be specified to necropsy personnel prior to the beginning of procedures.
- 35.2.8. Radioactive animal carcasses and tissues must be collected and stored for decay by staff trained in Radiation Safety.

RAD-030, RADIATION SAFETY MANUAL

35.2.9. Waste logs must be maintained. Radiation survey must be conducted after 10 half-lives, and survey results must be documented in the waste log. Radiation levels must not exceed background radiation levels.

35.3. Standard Protocols for Care of Animals Containing Radioactivity

35.3.1. Authorized radioisotope users who require care for animals treated with radioactive materials must provide, by direct supervision and/or complete written instructions, the procedures which the animal caretakers must follow with respect to cage handling and collection and disposal of radioactive waste.

35.3.2. Investigators often house animals containing radioactive material in general animal care facilities which are used by several investigators at the same time. Such facilities, by necessity, are accessible to people with widely varying training in radiation safety.

35.3.3. To assist investigators who must maintain animals treated with radioactive material, a set of three standard protocols have been developed for animal care and specify the radiation conditions permitted for their use. Animal care protocols must be specified in the IACUC protocol.

35.3.4. Each protocol is designated by a specific color card.

35.3.4.1. The color alerts any person in the vicinity of the animal cage to the presence and degree of potential radiation hazard.

35.3.4.2. The decision on whether animals containing radioactive material may be housed in general care facilities or if they must be placed in isolation rooms depends in part on the radiotoxicity of the radionuclides being used and the maximum activity excreted daily per cage or room. Limits are based on the chemical and physical forms of the radionuclide and its excretion rate.

35.3.5. Radiotoxicity classifications and maximum daily excretion for commonly used radionuclides are as follows:

Table 8, Radiotoxicity Classification

RADIOTOXICITY CLASSIFICATION	
Hazardous	Ca-45, I-123, I-125, I-131
Moderately Hazardous	Cr-51, Cu-64, Ga-68, F-18, P-32, S-35, Se-75, Tl-201
Slightly Hazardous	H-3, C-14, Tc-99m, In-111

RAD-030, RADIATION SAFETY MANUAL
Table 9, Maximum Permitted Daily Activity Excretion

MAXIMUM PERMITTED DAILY ACTIVITY EXCRETION			
Radiotoxicity Classification	HOUSED IN GENERAL ANIMAL FACILITIES		HOUSED IN ISOLATION ROOMS
	Care by Non- Radiation Workers (Blue Protocol)	Care by Radiation Workers (Yellow Protocol)	Care by Radiation Workers (Red Protocol)
Hazardous	< 100 μ Ci / cage and < 500 μ Ci / room	> 100 μ Ci /cage and < 500 μ Ci / room	> 500 μ Ci / room
Moderately Hazardous	< 1 mCi / cage and < 5 mCi / room	> 1 mCi / cage or > 5 mCi / room	> 25 mCi / room
Slightly Hazardous	< 5 mCi / cage and < 10 mCi / room	> 5 mCi / cage or > 10 mCi / room	> 50 mCi / room

36. GUIDELINES FOR SELECTING PROPER ANIMAL CARE PROTOCOL

Each of the standardized protocols for animal care is designated by a color coded card. Selection of appropriate protocol (and hence the appropriate cage card) is essential to ensuring the safe handling of animals containing radioactive material. Animal care protocols are selected based on radiotoxicity, isotope half-life, dose rate, and likelihood of contaminated bedding. In order of increasing potential hazard, the color codes used are as follows: Blue - Yellow - Red.

36.1. Blue Protocol – Used with any animal that does not require care by radiation workers or isolation.

- 36.1.1. Used only with short-lived isotopes (physical half-life < 10 hours) as long as the quantity does not exceed the limits established for non-radiation workers. Radiation dose rate in normal work space does not exceed 2 millirem/hour. Card may not be appropriate with animals requiring daily pen/cage cleaning unless the physical half-life of the radionuclide is less than 2.5 hours. Radiation dose rate in normal work space does not exceed 2 millirem/hour.
- 36.1.2. Researchers must complete the cage cards based on the information posted in the scan room. Cards will be returned to the Radiation Permit Holder's staff at the end of the study.

RAD-030, RADIATION SAFETY MANUAL

36.1.3. Animal holding area surveys must be conducted and documented on the card by research staff. Radiation levels for animal bedding and waste must be checked at the end of the study. Radiation levels must not exceed background radiation levels.

36.2. *Yellow Protocol - Used with any animal requiring care by radiation workers, but not requiring isolation.*

36.2.1. Used for radioisotopes with a physical half-life >10 hours, as long as the quantity does not exceed the limits established for non-radiation workers. The radiation dose rate in the immediate work area may exceed 2 millirem/hour, but the radiation dose rate in adjacent work areas must not exceed 2 millirem/hour.

36.2.2. Animal litter is contaminated and must be collected and stored as radioactive waste. The Radiation Permit Holder or other trained radiation worker provides all routine care of animal as required by DAR for the designated care period.

36.2.3. Records must identify the persons who provide animal care. Animal litter and waste will be collected by trained radiation worker. Waste must be checked before regular disposal.

36.2.4. Cage must be checked for contamination at the end of the study. Cage cards must be returned to Radiation Permit Holder's staff assigned for the area.

36.2.5. Animal holding area surveys must be conducted and documented on the card by the Radiation Permit Holder or other trained radiation worker. Radiation levels must not exceed background radiation levels.

36.3. *Red Protocol - Used for any animal requiring isolation.*

36.3.1. The degree of hazard requires isolation of the treated animal(s).

36.3.2. Access to the isolation room is restricted to necessary personnel.

36.3.3. Arrangements for an isolation room must be made with DAR prior to the start of the experiment.

36.3.4. The Radiation Permit Holder together with Radiation Safety and DAR will determine the duration of isolation and the specific procedures to maintain satisfactory animal care and radiation protection.

36.3.5. The Radiation Permit Holder or other trained radiation worker provides all routine care of the feeding, watering, cage washing, and room cleaning.

36.3.6. Waste and contaminated equipment must be collected and stored for decay by the Radiation Permit Holder or other trained radiation worker. The room must be checked by Radiation Safety staff before releasing room for general use.

RAD-030, RADIATION SAFETY MANUAL**INSTRUMENTATION****37. GENERAL INFORMATION****37.1. *What Instruments Do I Need To Work With Radioactivity?***

- 37.1.1. You are required to list the surveying and counting instrumentation available for use in your laboratory when applying for an authorization.
- 37.1.2. When an application enters the Radiation Safety Office, it is reviewed by a staff member who determines that the correct instrumentation is available to detect the radionuclides requested.
- 37.1.3. If a necessary piece of instrumentation is missing, the Radiation Permit Holder must purchase the equipment or receive permission from another Radiation Permit Holder to use their equipment.
- 37.1.4. Most laboratories purchase their own GM survey meter when needed, but liquid scintillation and gamma counters are often shared within departments.
- 37.1.5. The equipment must be on order prior to submission of the application to the appropriate committee.

37.2. *How Do I Obtain A GM Survey Meter?*

A GM survey meter can be purchased from Ludlum measurements, Johnson Nuclear, Thermo Scientific, or other major manufacturer of radiation detection equipment. The most popular and commonly used GM survey meter at Emory is the Ludlum Model 3 with a Model 44-9 “pancake” probe.

37.3. *What Are the Requirements Of GM Survey Meter and Survey Instruments?*

- 37.3.1. The survey meter must have an external detector for surveying surfaces for contamination and can read in either mR/hr, CPM, CPS or a combination.
- 37.3.2. For Clinical Use:
 - 37.3.2.1. For uptake, dilution & excretion studies for which a written directive is not required, a meter with a range of 0.1 to 50 mrem/hr is required;
 - 37.3.2.2. For procedures requiring a written directive, a meter with a range of 0.1 to 50 mrem/hr and a meter with a range of 1 mrem/hour to 1000 mrem/hour is required; and
 - 37.3.2.3. There must be a dedicated check source available.

37.4. *How Do I Get My GM Survey Meter Calibrated?*

Survey meters must be picked up by the Radiation Safety Office for calibration before first use (unless accompanied by a calibration certificate), annually, and following any repair that will affect the calibration. The calibration sticker on the instrument will indicate the date by which the instrument must be recalibrated. The record of each meter calibration must be maintained by the Radiation Permit Holder for three years or per Emory Policy, whichever is longest. It is the responsibility of the Radiation Permit Holder to see that the instrument is free of contamination. Most laboratories can arrange to share an instrument with a neighboring lab while their instrument is being calibrated, but some “loaner” meters are available.

RAD-030, RADIATION SAFETY MANUAL**37.5. *How Long Does Calibration Take?***

Instrument calibration can take up to 2 weeks but may take longer if additional repairs are needed.

37.6. *How Does the Meter Loaner Program Work?*

The meter loaner program is a service provided in the event your meter is due for calibration and you are actively working with radioactivity. A loaner will be provided until calibration or repair of your meter is completed and must be returned after service is completed.

37.7. *How Do I Get My GM Survey Meter Repaired?*

If your survey meter requires repairs contact Radiation Safety for an assessment. Radiation Safety may be able to perform some small repairs. Repairs outside the scope of our services must be completed by the instrument manufacturer at your expense. Instruments returned to the manufacturer must be wipe tested prior to shipment.

37.8. *QC Requirements for Gamma Counters/Well Counters/Liquid Scintillation Counters/Thyroid Probes*

37.8.1. Quality control checks for gamma counters, well counters, liquid scintillation counters, and thyroid probes must be performed as recommended by the manufacturer. These checks must include a background and constancy check each day the instrument is used for clinical purposes.

37.8.2. Periodic tests to assure proper performance of the instrument, such as a chi-square test, must be performed quarterly, or as recommended by the manufacturer.

37.8.3. The efficiency of the well counter for the isotopes of interest and the window of detection must be checked annually for non-human use, daily for human use.

37.8.4. Procedures and forms for the tests mentioned here are available from the Radiation Safety Office.

37.8.5. Records must be kept according to [Radiation Safety Office Record Keeping Guidelines.](#)

37.9. *What Dose Calibrator QC is Required?*

37.9.1. An Authorized User authorized to administer radiopharmaceuticals shall possess and use a dose calibrator to measure the activity of unsealed radioactive material prior to administration to each patient or human research subject.

37.9.2. The dose calibrator must be tested in accordance with nationally recognized standards or the manufacturer's instructions which shall include the following:

37.9.2.1. Constancy check each day of use with sealed source(s) of not less than 50 microcuries of a photon-emitting source, such as Cs-137.

37.9.2.2. Linearity check upon installation and quarterly thereafter and following adjustment or repairs.

37.9.2.3. Accuracy check annually and following adjustment or repair.

37.9.2.4. Geometry dependence upon installation and following adjustment or repair.

37.9.2.5. Records must be kept according to [Radiation Safety Office Record Keeping Guidelines.](#)

RAD-030, RADIATION SAFETY MANUAL**37.10. *How Do I Dispose Of Old Instruments Such As A Liquid Scintillation Counter?***

If you need to dispose of a liquid scintillation counter, gamma counter, or Geiger counter, contact Radiation Safety in order to test the instrument for contamination, check for internal radioactive sources, and properly dispose of the instrument.

Liquid scintillation counters contain Cs-137, Ba-133, or Ra-226. The instrument manufacturer must remove the radioactive source from the instrument before disposal.

37.11. *How Do I Order Instrument QC Sources?*

Order your instrument QC check sources using the [radioactive materials ordering guidelines](#).

37.12. *What Should I Do If My Survey or Counting Instruments Fail?*

If your survey or counting instrument fails and you are unable to survey for contamination, contact Radiation Safety to find a temporary replacement while your instruments are repaired.

37.13. *Other Radiation Equipment*

Area monitors, hand & foot monitors, and stack monitors must be calibrated on an annual basis.

Breathing zone monitors may be necessary if working with volatile radioactive compounds within a fume hood. The monitors will be calibrated annually and available upon request. A base line bioassay may be necessary depending on the radioisotope of interest.

RAD-030, RADIATION SAFETY MANUAL**USE OF RADIOACTIVE MATERIAL IN HUMANS**

The use of radioactive material in humans is specifically regulated in Chapter .05 of the Georgia regulations for radioactive material use (391-3-17). The following sections summarize these regulations and any applicable license conditions.

The user should also familiarize themselves with the responsibilities of the Authorized User (AU), the Authorized Medical Physicist (AMP), and the Nuclear Medicine technologist (NMT) in section 4 of this Manual.

38. WHO IS AUTHORIZED TO ADMINISTER RADIOACTIVE MATERIAL TO HUMANS?**38.1. *Can Any Physician Read Nuclear Studies or Prescribe Radioactive Material for Human Use?***

No. Physicians wishing to prescribe or direct the use of radioactive material to patients or research subjects or prepare or supervise the preparation of radioactive material for use in humans must first be approved by the Radiation Safety Committee 1 and designated on the Clinical Authorization as an Authorized User for the specific use of radioactive material.

The specific uses of radioactive material are categorized as follows:

- 38.1.1. Uptake, Dilution or Excretion Studies;
- 38.1.2. Imaging and Localization Studies;
- 38.1.3. Unsealed Use of Radioactive Material Requiring a Written Directive;
- 38.1.4. Manual Brachytherapy
- 38.1.5. Ophthalmic Use of Sr-90
- 38.1.6. Sealed Sources in a Remote Afterloader Unit

38.2. *How Do I Become an Authorized User?*

Physicians wishing to be listed as Authorized Users:

- 38.2.1. Must be a faculty member of Emory University;
- 38.2.2. Must submit the following documentation to the Radiation Safety Office:
 - 38.2.2.1. current State of Georgia Medical License;
 - 38.2.2.2. CV;
 - 38.2.2.3. Board Certification, if available;
 - 38.2.2.4. Completed, Signed Committee I Amendment Form
 - 38.2.2.5. and either
 - 38.2.2.5.1. A copy of a Radioactive Materials License from another institution that shows the physician listed as an Authorized User for the specific use of radioactive material for which a written directive is required; or

RAD-030, RADIATION SAFETY MANUAL

38.2.2.5.2. A preceptor form signed by an Authorized User who supervised his training and experience.

38.2.2.5.3. For Nuclear Cardiologists, use preceptor form nrc313a (aud);

38.2.2.5.4. For Nuclear Medicine, use preceptor form nrc313a (aut);

38.2.2.5.5. For Radiation Oncology, use preceptor form nrc313a (aus).

38.2.3. Must receive documented training in applicable Emory Radiation Safety policies and procedures.

38.2.3.1. Once the required documentations are received, they will be submitted to the Radiation Safety Committee 1 for review and approval. Once approved, the physician and their department will receive written notification that they may function as an Authorized User.

38.3. *Who Else Is Authorized To Administer Radioactive Material To Patients or Research Subjects?*

Only individuals designated in writing by an Authorized User and trained by the Radiation Safety Office are permitted to prepare or administer radioactive material to patients. These individuals must be listed on the Clinical radioactive material Authorization.

Please contact the Radiation Safety Office to schedule Radiation Safety training, or if you need assistance with the qualification documentation requirements.

38.4. *New Nuclear Medicine Technologists*

A staff member that wishes to function as a Nuclear Medicine Technologist, prior to their first use of radioactive material, must:

38.4.1. Submit the following to the Radiation Safety Office:

38.4.2. Completed, Signed RSC 1 Amendment Form; and

38.4.3. Copy of their Certification in Nuclear Medicine Technology (or documentation of other qualification as stipulated in Georgia rule 391-3-17.05(25));

38.4.4. Obtain written designation by an Authorized User that the person may procure and inject radiopharmaceuticals in patients or research subjects;

38.4.5. Receive documented training in applicable Emory Radiation Safety policies and procedures.

38.5. *New Authorized Medical Physicists*

An Authorized Medical Physicist must be approved by the RSC1 and designated on the Clinical Authorization prior to their first use of radioactive material:

38.5.1. Submit the following to the Radiation Safety Office:

38.5.1.1. Completed, Signed RSC 1 Amendment Form; and

38.5.1.2. Copy of their Board Certification (or documentation of other qualification as stipulated in Georgia rule 391-3-17.05(23) or (26));

38.5.2. Receive documented training in applicable Emory Radiation Safety policies and procedures.

RAD-030, RADIATION SAFETY MANUAL**39. GENERAL REQUIREMENTS FOR CLINICAL USE OF RADIOACTIVE MATERIAL****39.1. Use of Protective Equipment**

Since potential occupational exposure from radioactive material used in clinical settings are generally much greater than laboratory settings, the following rules are established to minimize occupational exposure.

- 39.1.1. Use syringe shields for reconstitution of radiopharmaceutical kits and administration of radiopharmaceuticals to patients, except in those circumstances in which their use is contraindicated (e.g., recessed veins, infants). In these exceptional cases, consider the use of other protective methods such as remote delivery of the dose (e.g., through use of a butterfly needle.)
- 39.1.2. Always use vial shields when preparing or handling a vial that contains a radiopharmaceutical.
- 39.1.3. Syringes that contain radioactive material must be kept in shielded containers that are clearly labeled.
- 39.1.4. Always keep flood sources, syringes, waste and other radioactive material in shielded containers.
- 39.1.5. Consider using a cart to move flood sources, waste, and other radioactive material, since sources with even small amounts of radioactivity exhibit a high dose rate on contact.
- 39.1.6. Authorized users must secure radioactive material, both in storage and in use, from unauthorized removal or access.
- 39.1.7. Departments must possess appropriate survey instruments. (See “What Are the Requirements Of GM Survey Meter and Survey Instruments?”)

39.2. What are the Rules for Using Gases, Aerosols and Volatiles?

Clinical departments that administer radioactive aerosols or gases shall:

- 39.2.1. Store such volatile radioactive materials and radioactive gases in the shippers' radiation shield and container.
- 39.2.2. Use and store multi-dose containers in a properly functioning fume hood.
- 39.2.3. Administer the gas/aerosol/volatile with a system that will keep airborne concentrations within regulatory limits. The system shall either be directly vented to the atmosphere through an air exhaust or provide for collection and decay or disposal of the aerosol or gas in a shielded container.
- 39.2.4. Check the operation of collection systems monthly. Records must be kept according to [Radiation Safety Office Record Keeping Guidelines](#).

39.3. Rules for Safe Dose Administration

The following rules are established to assure that radioactive material is administered to a patient as directed by the Authorized User:

- 39.3.1. Radiation may only be administered to patients according to a written directive or by reference to the diagnostic clinical procedures manual.

RAD-030, RADIATION SAFETY MANUAL

39.3.2. Radiopharmaceutical multi-dose diagnostic and therapy vials must be labeled with the radionuclide, the activity, the date for which the activity is calibrated, and the radiopharmaceutical.

NOTE: Radiopharmaceuticals used under RDRC require additional labeling. See RDRC Labeling SOP.

39.3.3. Syringes and unit doses must be labeled with the radioactive drug.

39.3.4. For prepared doses, patient doses must be assayed before administration either in the dose calibrator or per manufacturer's instructions.

39.3.5. Do not use a dose that varies by greater than $\pm 20\%$ from the prescribed dose, except for prescribed doses of less than 30 microcuries or as approved by the authorized user,

39.3.6. When measuring the dose, the amount of radioactivity that adheres to the syringe wall or remains in the needle need not be considered.

39.3.7. Confirm the patient's name and identification number and the prescribed radionuclide, chemical form, and dose before administering.

39.3.8. In the event that a dose calibrator is malfunctioning, the manufacturer's assay can be used by decay correction and volumetric assay.

39.3.9. If the prescribed dose requires a written directive, the patient's identity must be verified and the administration must be in accordance with the written directive.

39.4. *Special Consideration for the Pregnant or Breastfeeding Patient*

If the patient is pregnant or breastfeeding, then special considerations must be given regarding the necessity of the procedure and the radiation risk to the fetus or the nursing child. Any administration that results in a dose of 500 mrem to a fetus or nursing child must be specifically approved in advance by the Authorized User. Also see [Release of Patients Containing Radioactive Material](#).

39.4.1. Additional safety precautions may be instituted by Authorized User or Radiation Safety Office, as needed.

39.4.2. Follow the rules in "[Patient Care](#)" for release of patients carrying radioactive material.

39.5. *What Information Must Dose Administration Records Contain?*

39.5.1. For all administrations record:

39.5.1.1. the patient's name or identification number,

39.5.1.2. prescribed dose (or procedure name in which the dose can be referenced from the Clinical Procedures Manual),

39.5.1.3. the determined dose (if > 30 uCi),

39.5.1.4. the date and time of dose determination,

39.5.1.5. The name of the individual who determined the dose.

39.5.2. If any administration error is discovered, determine the patient's name, administered dose, pharmaceutical, date, time and route of administration and report to the Authorized User, the Radiation Safety Office and the SAFE reporting system.

RAD-030, RADIATION SAFETY MANUAL**40. GUIDELINES FOR WRITTEN DIRECTIVES (PRESCRIPTIONS)**

A “written directive” is a written order by an Authorized User for the administration of radioactive material to a specific patient or human research subject.

40.1. Which Administrations Require a Written Directive?

40.1.1. A written directive must be prepared for:

- 40.1.1.1. any administration of I-131 sodium iodide greater than 1.11 MBq (30 uCi),
- 40.1.1.2. any therapeutic dose of a radiopharmaceutical, and
- 40.1.1.3. all therapeutic doses of radiation from byproduct material (e.g., manual and high dose rate remote afterloading brachytherapies)
- 40.1.1.4. Administrations of radioactive material or radiation from radioactive material using radioactive material approved for routine use by the FDA that are beyond clinically indicated uses,
- 40.1.1.5. Administrations of radioactive material or radiation from radioactive material that use radioactive material approved under an Investigational New Drug or expedited Investigational New Drug (IND, eIND) or by the RDRC.

40.2. What Information Must Be On the Written Directive?

40.2.1. Written directives must contain the patient or research-subject’s name and the following:

- 40.2.1.1. For an administration of a dosage of radioactive drug containing radioactive material, the radioactive drug containing radioactive material, dosage, and route of administration;
- 40.2.1.2. For high dose rate remote afterloading brachytherapy, the radionuclide, treatment site, dose per fraction, number of fractions, applicator used and total dose;
- 40.2.1.3. For all other brachytherapy
- 40.2.1.4. Prior to implantation: treatment site, the radionuclide, and dose; and
- 40.2.1.5. After implantation but prior to completion of the procedure: the radioisotope, treatment site, number of sources, and total source strength and exposure time (or, the total dose).
- 40.2.1.6. Signature of an Authorized User approved for the modality of the administration.

40.3. Written Directive Procedures

40.3.1. Departments must develop, implement and maintain written procedures for administrations requiring written directives that address, as applicable:

- 40.3.1.1. Verifying the identity of the patient or human research subject;
- 40.3.1.2. Verifying that the specific details of the administration are in accordance with the treatment plan, if applicable, and the written directive;
- 40.3.1.3. Checking both manual and computer-generated dose calculations; and
- 40.3.1.4. Verifying that any computer-generated dose calculations are correctly transferred into the HDR console.
- 40.3.1.5. Any dose calculations are checked and verified

RAD-030, RADIATION SAFETY MANUAL

- 40.3.1.6. Prior to administering a dose, the patient's or human research subject's identity will be verified as the individual named in the Written Directive.
Examples of patient identity verification include the patient's I.D. bracelet, hospital I.D. card, driver's license or social security card,
- 40.3.1.7. Components of the Written Directive (radionuclide, total dose, route of administration) will be confirmed by the person administering the dose to verify agreement with the Written Directive,
- 40.3.2. A Call to Order is required for all administrations to females requiring a written directive. The Call to Order consists of the verification of the results of a pregnancy test and must be included in the prescription checklist. The Call to Order must be conducted by:
- 40.3.2.1. a technologist and an attending physician, or
- 40.3.2.2. A resident and an attending physician.
- 40.3.3. Record the radiopharmaceutical dose or radiation dose actually administered.
- 40.3.3.1. An Authorized User must date and sign a written directive prior to the administration of any dose,
- 40.3.3.2. A prescription is required for administrations of radioactive material or radiation from radioactive material that are beyond standard of care, using radioactive material approved for routine use by the FDA. Any Authorized User approved for that modality may sign the prescription.
- 40.3.3.3. A prescription is required for administrations of radioactive material or radiation from radioactive material that are beyond standard of care, that use radioactive material approved under an IND, eIND or by the RDRC. The prescription shall be signed by an Authorized User approved for that modality, who is listed on the IRB protocol.¹
- 40.3.3.3.1.1. The PI will identify the Authorized Users listed on the protocol for each approved protocol in category.
- 40.3.3.3.1.2. The PI will report any changes to the list to Radiation Safety.
- 40.3.3.3.1.3. Copies of Written Directive shall be maintained for audit purposes.

41. BRACHYTHERAPY RULES

In this section and Section 45, references to “GA. Rule .05##” refer to Georgia Department of Natural Resources’ Rules and Regulations for the Use of Radioactive Materials in the Healing Arts, Chapter 391-3-17-.05.

41.1. Technical Requirements

- 41.1.1. Only sources approved in Sealed Source Device Registry or with an Investigational Device Exemption can be used.
- 41.1.2. Must follow the manufacturer’s safety & handling instructions.

¹ For studies under an IND or eIND the prescription should be signed by an individual listed on Form FDA 1572

RAD-030, RADIATION SAFETY MANUAL

41.1.3. Must have an appropriate survey instrument. (See Instrumentation Section “What Are the Requirements Of GM Survey Meter and Survey Instruments?”)

41.2. Determining Brachytherapy Source Activity:

41.2.1. Source output/activity must be determined before first use using an acceptable dosimetry system (manufacturer’s measurements are acceptable);

41.2.2. Positioning accuracy must be determined within applicators;

41.2.3. Source activity must be decay corrected activity at 1% decay intervals;

41.2.4. Retain records of these calibrations and decay calculations according to [Radiation Safety Office Record Keeping Guidelines.](#)

41.3. Acceptance Testing

Acceptance testing of computer systems must be performed in accordance with published protocols. At a minimum, testing should meet requirements in GA. Rule .05(61).

41.4. Brachytherapy Procedures

41.4.1. Must follow Written Directive Procedures above,

41.4.2. Must maintain brachytherapy source accountability at all times (see Inventory below)

41.4.3. Patients must be surveyed after implant and removal

41.4.4. Department must take emergency response equipment to procedure room to respond to a source that becomes inadvertently dislodged from the patient.

41.5. Specific Nursing precautions:

See also Patient Care [Section](#)

41.5.1. For Cs-137 tandem and ovoid and Ir-192 interstitial implants:

41.5.1.1. Patients are not allowed to leave the room until cleared by the Physicist

41.5.1.2. Nothing leaves the room until cleared by physicist

41.5.1.3. Dietary and housekeeping are not allowed into the room.

41.5.1.4. Lab is not allowed in the room without consulting Radiation Oncology or Radiation Safety Office.

41.5.2. For patients receiving eye-plaques:

41.5.2.1. No additional precautions beyond normal nursing care procedures.

41.6. Brachytherapy Inventory

41.6.1. The department shall maintain accountability at all times for all brachytherapy sources in storage or use.

41.6.2. An inventory record must be maintained that includes all items required by GA Rule.05 (103).

41.6.3. Use the EHS Assistant database to account for any sources that are to be sent to the Radiation Safety Office for decay-in-storage

RAD-030, RADIATION SAFETY MANUAL**42. HIGH DOSE RATE (HDR) AFTERLOADER RULES****42.1. Installation, Maintenance, Adjustment, & Repair**

- 42.1.1. For any source installation or any repair of the high dose rate afterloader (HDR) unit, follow GA Rule .05(69)
- 42.1.2. Reciprocity of license must be in effect when handling radioactive sources at Emory locations
- 42.1.3. Following any new source installation or unit repair, surveys must be performed according to GA Rule .05 (80)
- 42.1.4. The facility must meet the requirements for interlocks and monitoring systems in GA Rule .05(71).
- 42.1.5. Acceptance testing of the treatment planning system must meet the requirements of GA Rule .05(82).

42.2. HDR Procedures

- 42.2.1. Only procedures which allow for expeditious removal of a decoupled or jammed source are permitted.
- 42.2.2. An Authorized User and an Authorized Medical Physicist must be physically present during all patient treatments involving the unit. During continuation of treatment, a physician, under the supervision of an Authorized User, who has been trained in the operation and emergency response for the unit, may replace the Authorized User.
- 42.2.3. No person is allowed to be present in the room with the patient during treatment, unless approved by the Authorized User, Radiation Safety Office, or Authorized Medical Physicist.
- 42.2.4. Emergency equipment must be available to respond to an unshielded source or a source lodged in the patient.
- 42.2.5. The department shall possess a portable radiation detection survey instrument capable of detecting dose rates over the range of 0.1 mrem/hr to 50 mrem/hr, and a portable radiation measurement survey instrument capable of measuring dose rates over the range 1 mrem/hr to 1,000 mrem/hr.

42.3. Safety Procedures

- 42.3.1. After the completion of a procedure, before they are released, the patient and the HDR unit must be surveyed with a portable radiation detection instrument to make sure that the source has returned safely to the unit.
- 42.3.2. The unit, the console, and the console keys must be secured when not in use or when unattended. Securing the unit within its locked enclosure is equivalent to securing the treatment room door.
- 42.3.3. Any person entering the HDR room following a procedure is required to use radiation monitors to verify that the radiation levels are safe.
- 42.3.4. Only one radiation producing device may be operated at a time in the room.
- 42.3.5. The department must have written procedures available for responding to abnormal situations (e.g., equipment failures, unable to return the source, etc.). See GA Rule .05(70) for required

RAD-030, RADIATION SAFETY MANUAL

content and posting of these procedures. Notify the Radiation Safety Officer, or his or her designee, and an authorized user as soon as possible, if the patient or human research subject has a medical emergency and, immediately, if the patient dies.

42.4. Required Training

- 42.4.1. All persons authorized to operate the unit must receive training in emergency situations (see previous rule above) and operating procedures.
- 42.4.2. AMPs, Authorized Users and operators must participate in emergency drills initially and annually.
- 42.4.3. Training records will be kept by Radiation Oncology according to [Radiation Safety Office Record Keeping Guidelines](#).

42.5. Full-Calibration Measurements Requirements

- 42.5.1. The department must use a dosimetry system that meets the requirements in GA Rule .05(72).
- 42.5.2. Full-calibration measurements must be performed by an Authorized Medical Physicist on the HDR before first use, following replacement of the source, or repair of the unit that includes removal of the source or major repair of the components associated with the source exposure assembly.
- 42.5.3. Full calibration measurements must be done once a quarter and must meet the requirements listed in GA Rule .05(74).

42.6. Spot Checks

- 42.6.1. Spot checks must be performed at the beginning of each day that the unit is used. The results must be documented.
- 42.6.2. The Authorized Medical Physicist must establish written procedures for the spot-checks. These procedures must meet the requirements in GA Rule .05(77)(d).
- 42.6.3. The Authorized Medical Physicist must review the results of the spot-checks within 15 days. Results only need to be reported to the Radiation Safety Office if any spot-check fails.
- 42.6.4. If the spot-checks indicate the malfunction of any system, a licensee shall lock the control console in the off position and not use the unit except as may be necessary to repair, replace, or check the malfunctioning system.

43. PATIENT CARE**43.1. Release of Patients Containing Radioactive Material**

Patients receiving diagnostic quantities of radioactive material are considered to be authorized for release. For patients receiving therapeutic amounts of radioactive material, the following rules apply:

- 43.1.1. An Authorized User may authorize the release of any individual who has received radioactive drugs or implants containing radioactive material if the total effective dose equivalent to any other individual from exposure to the released individual is not likely to exceed 5 mSv (0.5 rem).
- 43.1.2. Release of the patient must be approved by an individual listed as an Authorized User on the departmental authorization for the type of radioactive material use of which the patient being

RAD-030, RADIATION SAFETY MANUAL

released has received. For radioiodine therapies, the final determination is based on responses to a patient questionnaire completed by the patient concerning his living conditions and other persons in the home, and on the opinion of the physician that the patient is competent to carry out instructions.

- 43.1.3. The released individual, or the individual's parent or guardian, will be provided with instructions, including oral and written instructions, on actions recommended to maintain doses to other individuals as low as is reasonably achievable.
- 43.1.4. If a breast-feeding infant or child could receive a radiation dose as a result of the release of the patient, the instructions shall include guidance on interruption or discontinuation of breast feeding and information of the potential consequences, if any, of failure to follow the guidance.
- 43.1.5. The Authorized User shall maintain a record of the basis for authorizing the release of an individual and of the instruction provided to breast-feeding women [Radiation Safety Office Record Keeping Guidelines](#).

44. NURSING CARE OF PATIENTS CONTAINING THERAPEUTIC AMOUNTS OF RADIOACTIVE MATERIAL

44.1. Instructions for Nurses:

- 44.1.1. Personnel caring for patients who are not authorized for release according to Release of Patients Containing Radioactive Material must receive radiation safety instruction initially and at least annually which includes patient control, visitor control, contamination control and waste control.
- 44.1.2. Specific Guidelines for Nurses caring for patients containing radioactive material can be found in the document "Radioactive Materials: Care for Patient Containing" which is available in Lotus Notes in the Emory Hospitals Policies database.

44.2. Inpatient Radiation Controls

- 44.2.1. The patient should be placed, on a nursing unit that has received radiation safety training.
 - 44.2.1.1. At Emory University Hospital, 8E is used for Bexxar therapies; 6G and HG are used for other radioiodine therapy, 9E for brachytherapy of the head and neck; 11 E for brachytherapy OB/Gyn.
 - 44.2.1.2. At EUH Midtown, Unit 71 is used for all radiation inpatients.
 - 44.2.1.3. Check with the hospital Radiation Safety Officer for any changes.
- 44.2.2. The patient must be placed in a private room with a private bath.
- 44.2.3. All superfluous equipment and supplies must be removed from the patient room. Remaining items and portions of the floor must be covered with plastic or some other waterproof material to prevent surface contamination and spread of contamination.
- 44.2.4. The patient's or human research subject's room shall be posted with a "Caution: Radioactive Material" sign and a note shall be posted on door or in patient's chart stating how long and where a staff member or a visitor may stay in the room.
- 44.2.5. The Radiation Safety Officer, or his/her designee and the Authorized User shall be notified immediately if the hospitalized patient dies or has a medical emergency.

RAD-030, RADIATION SAFETY MANUAL**45. REPORTING PATIENT MEDICAL EVENTS**

Misadministrations are specifically defined in GA Rule 391-3-17-.05 (115). These definitions involve dose thresholds for whole body or organs. Instead of independently determining whether or not a misadministration has occurred, a department should report any unusual incident to the Radiation Safety Office. The Radiation Safety Office can then assist in determining if the incident qualifies as a reportable misadministration and, if so, reporting the incident to the proper agencies.

NOTE: Extravasated diagnostic doses are not considered misadministrations.

45.1. Incidents to Report

The following incidents shall be reported to the Authorized User and the Radiation Safety Office immediately upon discovery:

45.1.1. Any administration of radioactive material to a patient that involves:

45.1.1.1. The wrong patient;

45.1.1.2. The wrong radiopharmaceutical;

45.1.1.3. An administered dose that is 20% or more different than the prescribed dose; or

45.1.1.4. An unintended dose to a pregnant woman or nursing child; or

45.1.2. For therapies:

45.1.2.1. The wrong route of administration; or

45.1.2.2. The wrong mode of treatment; or

45.1.3. For brachytherapies:

45.1.3.1. A dose to tissue other than the treatment site that exceeds 50% of the expected dose defined in the written directive (excluding prostate seed migration), or

45.1.3.2. A leaking sealed source.

45.1.4. All such incidents will be investigated by the Radiation Safety Office to determine the cause and any actions that can be taken to prevent a recurrence.

45.1.5. If an incident is determined to be a misadministration, the Authorized User will assist the Radiation Safety Office in writing a description of the incident and in gathering the information required for notifying the DNR or any other agency. The referring physician must be contacted within 24 hours. The referring physician must decide who will notify the patient, or decide that notifying the patient would be harmful. Appropriate medical care, including remedial care from the misadministration, must not be delayed due to notification delays. The patient (or guardian) must be informed of the availability of the written description of the incident and provided a copy if requested. Reporting, patient notification and investigations must be in accordance with GA Rule .05 (115).

RAD-030, RADIATION SAFETY MANUAL**AUTHORIZATION INSPECTIONS**

The Radiation Safety Office visits each active area of radioactive material use quarterly in order to maintain safety and compliance with rules, regulations and license conditions. It is important for all users to understand that the ability to use radioactive material is a privilege and also a responsibility. The goal of the Radiation Safety Office inspection is to maintain that privilege. Thus, the inspector makes an effort to discuss findings with the lab worker to elevate their understanding of those responsibilities.

46. LABORATORY AUDITS**46.1. *Inspection Procedure***


In preparation for inspecting a laboratory, the research radiation safety liaison reviews the authorization for laboratory location; name of laboratory contact; names and training status of laboratory employees; possession limits; amount of each radionuclide on hand with transaction histories; and manufacturer, model and calibration due date of any survey meters.

The inspector uses the Emory Lab Safety Inspection criteria to determine laboratory compliance with applicable standards. Answers to checklist items are gathered through observation, conversation and review of records that are checked for presence and timely completion. Inspectors can provide performance-based instruction to users and technical level staff while observing their work practices.

GM surveys and wipe tests (as appropriate) are conducted in each laboratory facility in which radioactive material is used. These results are documented consistent with the requirements for survey records (see below). In addition to the surveys performed by Radiation Safety, each laboratory must be surveyed by the occupants each week that radioactive material is used in the laboratory.

46.2. *Keys to a Successful Inspection:*

- 46.2.1. Area surveys – GM survey and wipe tests performed within required time & recorded correctly
- 46.2.2. Database – updates made as soon as possible but no later than within 2 weeks of isotope use, accurate records, use, and disposal data
- 46.2.3. Isotope use – logs available, accurately maintained, kept on file
- 46.2.4. Notebooks - organized, current, complete
- 46.2.5. Authorization – all users authorized and trained, labs authorized
- 46.2.6. Dosimetry – worn correctly, stored properly, returned timely, and reports posted
- 46.2.7. Equipment – GM survey meters calibrated annually, and operational
- 46.2.8. Attire – lab coats, disposable gloves, goggles available, closed toe shoes
- 46.2.9. Laboratory – local shielding, no food or drink consumed or stored
- 46.2.10. Storage – secure
- 46.2.11. Waste – proper containers, not overfilled, labeled, shielded if needed
- 46.2.12. Postings – door postings, emergency procedures & phone numbers
- 46.2.13. Animals – labeled, solid cage bottoms, appropriately color coded.

 EMORY UNIVERSITY	Environmental Health and Safety Office Research Administration	Effective Date: March 8, 2017	Version: 7	Page: 72 of 80
RAD-030, RADIATION SAFETY MANUAL				

46.3. *Reports*

A detailed report is prepared itemizing the inspector's findings and sent to the Radiation Permit Holder. A scorecard system with color codes for deficiencies will be followed. For quarterly inspections, a zero-deficiency is green, a 1-item deficiency is yellow, and 2 or more deficiencies is red.

46.4. *Repeat Violations*

If serious deficiencies which have a high potential for resulting in excessive radiation exposure leading to possible personal injury are found, the Radiation Permit Holder is notified immediately of the situation. Lack of corrective action on the part of the Radiation Permit Holder can result in suspension of operation requiring action by the Radiation Control Council for reinstatement. Serious or repeated deficiencies may require the RPH to come before RSC2 and/or the Radiation Control Council to address the matter.

46.5. *Corrective Actions*

Radiation Permit Holders are required to take corrective actions and instruct lab personnel that they are aware of the deficiency and the actions being taken to correct the deficiency.

47. CLINICAL AUDITS

47.1. *Inspection Procedure*

In preparation for inspecting a clinical department, the health physicist reviews the Authorization for laboratory location; name of laboratory contact; names and training of employees authorized to handle radioactive material; history of inspection categories from previous inspections; authorized use of radioactive material; sealed source inventory and leak-test dates; and QC due dates of any survey meters, counters or other instruments used to detect or measure radioactive material. Most inspections are unannounced unless the department is off-site or seldom staffed.

The inspector uses the Emory Radiation Safety Inspection criteria appropriate for the type of clinical use of radioactive material to determine laboratory compliance with applicable standards. Answers to checklist items are gathered through observation, conversation and review of records that are checked for presence and timely completion. Any items of concern found during the inspection will be discussed during the inspection with the technologist and/or Authorized User, if available. Inspectors can provide performance-based instruction to users and technical level staff while observing their work practices.

G-M surveys (when appropriate) and wipe tests are conducted in areas where radioactive material is used or stored. These results are documented consistent with the requirements for survey records below.

47.2. *Reports*

A cover letter is prepared for each report detailing the inspector's findings. The report is sent to both the responsible Authorized User, to whom the Authorization belongs, and the Chief Nuclear Medicine Technologist or Medical Physicist.

47.3. *Corrective Actions*

For items of serious concern a written Notice of Violation will be issued. Examples of such items include but are not limited to; incidents which are in direct violation of regulations or license conditions, or acts or omissions that have a high potential for resulting in excessive radiation exposure leading to

RAD-030, RADIATION SAFETY MANUAL

possible personal injury. The department in violation is then required to submit to Radiation Safety corrective actions in writing to correct the deficiency. If the violation continues to be repeated in subsequent inspections, the Radiation Safety Office may call the appropriate personnel in the department to come before Radiation Safety Committee 1 and/or the Radiation Control Council to address the matter.

47.4. Surveys by Radiation Safety

Quarterly dose rate measurements and surface contamination checks will be performed by research radiation safety liaison in labs with radioactive material inventory. The survey by Radiation Safety does not relieve the authorized individual from the responsibility of conducting and documenting their own surveys.

47.4.1. If all results are negative, report will be filed without notification of the laboratory.

47.4.2. Positive results initiate a phone call to the laboratory radiation safety representative or Radiation Permit Holder. The situation must be resolved by documented cleaning and confirmatory wipes.

48. RADIATION SAFETY OFFICE AUDIT RECORDS

Records of all audit material will be maintained for inspection by the State Radioactive Materials Program according to [Radiation Safety Office Record Keeping Guidelines](#). The records will include date of audit, name of person conducting the audit, persons contacted by the auditor, areas audited, audit findings and corrective actions..

RAD-030, RADIATION SAFETY MANUAL**MACHINE-PRODUCED RADIATION**

The term “radiation-producing machines” includes diagnostic x-ray machines, analytic x-ray machines, x-ray diffraction devices, DEXA machines, cabinet x-ray machines, cyclotrons, linear accelerators and electron microscopes. All radiation-producing machines operated at Emory University and associated institutions are operated under the jurisdiction of the State of Georgia, Department of Community Health. All uses of machine-generated radiation are to be carried out in accordance with the State of Georgia Rules and Regulations for X-Ray, Chapter 290-5-22.

49. REQUIREMENTS FOR POSSESSION OF RADIATION-PRODUCING EQUIPMENT**49.1. Registration**

- 49.1.1. All machines and devices designed to produce x-ray or which produce x-rays incidental to their operation shall be registered with the State of Georgia Department of Community Health.
- 49.1.2. Notify the Radiation Safety Office of the make, model, serial number and location of any new radiation producing device.
- 49.1.3. The Radiation Safety Office shall be notified when there is any change in the setup of the unit. Such changes include new equipment installed; changes in shielding of the surrounding walls; repair resulting in change in output of radiation; or change in usage of the unit.

49.2. Posting

- 49.2.1. Areas in which radiation-producing machines are located or are being used shall be posted with the characteristic "Caution Radiation" sign.

EXCEPTION: Diagnostic and patient treatment areas need not be so marked, provided that a person is charged with the responsibility for protection of employees, patients, and authorized visitors against unnecessary radiation and for the execution of Radiation Safety recommendations.

- 49.2.2. In addition, the controls shall bear a decal with the statement: "WARNING: This x-ray unit may be dangerous to patient and operator unless safe exposure factors and operating instructions are observed."

49.3. Surveys

- 49.3.1. All protective devices that may become defective due to use or abuse, such as protective lead aprons, skirts, vests, thyroid collars, and gloves, should be inspected for radiation leakage at least annually, and whenever the integrity of the equipment is suspect.
- 49.3.2. An annual, scheduled survey of all radiation-producing equipment used on patients shall be made by Radiation Safety personnel or a qualified expert. In addition, radiation surveys will be made of all new installations and all existing installations after every change that might affect the radiation output (e.g., replacement of x-ray tube, changes of filtration in beam).
- 49.3.3. Entrance exposure rates for the beam shall be determined for all units used on human subjects and must be made available to the operator by posting this information on the equipment or in the control room.

RAD-030, RADIATION SAFETY MANUAL

49.3.4. The faculty owners of radiation producing devices will ensure that each device under their control is surveyed annually by Radiation Safety personnel or a qualified expert to ensure that its performance is within legal requirements and that it is functioning correctly and safely.

49.4. *Shielding for X-Ray Machines*

49.4.1. With the exception of dental, bone density devices and mammographic machines, all X-Ray machines designated to be used within a specific location must have the walls of that location shielded to contain the radiation and reduce the exposure of those outside of the X-Ray room to within legally designated levels.

49.4.2. The structural shielding requirements of any new installation, or an existing one in which changes are contemplated, shall be approved by the Radiation Safety Office or a qualified expert and the Ga. Department of Community Health.

49.4.3. A radiation shielding integrity survey will be performed on all new or modified x-ray rooms by Radiation Safety personnel or a qualified expert.

50. POLICIES FOR STAFF**50.1. *General Policies for Safe Use of X-Ray Equipment***

All operators of portable x-ray equipment, or those who are likely to be exposed in one year to 10% of the occupational dose limits (see [Table 1](#)) must request and wear appropriate exposure monitoring devices.

50.1.1. Operators who are issued dual dosimeters must wear the body dosimeter beneath the lead apron at the waist or chest level and wear the collar dosimeter outside the lead apron.

50.1.2. Operators who are issued a single dosimeter must wear the dosimeter outside their lead apron, at the collar level.

50.1.3. The operator must keep exposures as low as reasonably achievable (ALARA) and must use minimum exposure factors necessary of the exam being performed. Fluoroscopic work shall be performed in the minimum time possible using the lowest dose rate and the smallest aperture consistent with clinical requirements.

50.1.4. The operator must never expose himself to the direct beam, and must not stand within one meter of the tube or irradiation target while the unit is in operation unless adequately shielded. The operator must make full use of protective barriers, lead aprons, gloves and goggles when practical.

50.1.5. The hand of the fluoroscopist should never be placed in the useful beam unless the beam is attenuated by the patient and a protective leaded glove is worn.

50.1.6. During the operation of mobile and dental units, the operator should stand as far as possible from the tube and patient during exposure, and should wear a protective apron, or step behind an adequate shield. Rotation of operators or the use of portable shields is recommended for heavy workloads.

50.1.7. Shutter mechanisms and interlocking devices should not be tampered with and must be inspected at frequent intervals to insure proper operation.

RAD-030, RADIATION SAFETY MANUAL

- 50.1.8. The operator should insist that all nonessential personnel leave the exposure area before operating the unit and that all essential personnel be adequately shielded.
- 50.1.9. The operator must observe any restrictions in the use of the unit recommended by the Radiation Safety Office staff.
- 50.1.10. The operator must notify their supervisor and the Radiation Safety Officer immediately of any accidental exposure to radiation to staff.
- 50.1.11. Use mechanical means to hold patients, animals or image receptors when necessary. Only when mechanical means are unusable should an employee hold the patient, animal or image receptor. No person shall be regularly employed to hold patients, animals or image receptors during exposure. The person holding the patient, animal or image receptor shall wear protective gloves and a protective apron. No part of this person's body should be in the unattenuated useful beam.

50.2. Who Can Operate X-Ray Equipment?

- 50.2.1. Only physicians and other licensed practitioners of the Healing Arts (as defined in the Rules and Regulations of the State of Georgia 290-5-22-.01(ff) and registered radiologic technologists under the direction of a physician, or persons with adequate training as described below under the direction of a physician, shall be allowed to apply x-rays from a machine to patients. Training courses and instructions for accessing them are listed on the [training page of the EHSO website](#).
- 50.2.2. Physicians will receive appropriate radiation safety [training](#) at orientation unless a formal course in radiation physics and safety occurred during training (residency, fellowship or as a faculty member). Alternatively, an X-ray Operators initial module will be taken at the time of requesting privileges. Fluoroscopy training is described in section 51.2.4.
- 50.2.3. Radiologic technologists must be graduates of an accredited program in radiologic technology and must be a member in good standing of the American Registry of Radiologic Technologists or be Registry eligible.

50.2.3.1. Radiologic technologists must comply with the Continuing Education Requirements for Renewal of Registration.

50.2.3.2. Radiologic technologists will receive [training](#) provided by EHSO.

50.2.4. Fluoroscopy Use

- 50.2.4.1. Physicians will receive training in fluoroscopic safety provided by the Credentialing Modules at initial credentialing and recredentialing, orientation unless a formal course in radiation physics and safety occurred during training (residency, fellowship, or as a faculty member) (see <http://www.emoryhealthcare.org/for-physicians/index.html>). Physicians who regularly operate a fluoroscope at the time this policy is implemented receive the above course upon recredentialing and all subsequent recredentialing.
- 50.2.4.2. Residents who are identified as fluoroscopy users receive [training](#) in fluoroscopic safety provided by EHSO and should be directly supervised (i.e., available in adjacent area) by a physician credentialed to perform fluoroscopy.
- 50.2.4.3. Fellows who are identified as fluoroscopy users receive [training](#) in fluoroscopic safety provided by EHSO.
- 50.2.4.4. Advanced Practice Professionals who perform fluoroscopy: Must be directly supervised by a physician credentialed to perform fluoroscopy; and: will receive [training](#) in fluoroscopic safety provided by EHSO.

RAD-030, RADIATION SAFETY MANUAL

50.2.5. The Radiation Control Council may recommend additional training if there are concerns about patient or employee radiation safety. Reasons for additional training could include, but are not limited to: (a) adoption of new fluoroscopic technology (b) high personal dosimetry badge readings that approach or exceed regulatory limits (c) observation of poor radiation protection practices or (d) resumption of fluoroscopic procedures after a long period of inactivity.

50.2.6. Other operators have courses available on the [training page of the EHSO website](#) specific to the hazards and regulations of their job function

50.2.6.1. DEXA operators must work at the direction of a physician and receive [training](#) specific to their job function.

50.2.6.2. Dentists will receive [training](#) specific to their job function.

50.2.6.3. Operators of dental units must work at the direction of a dentist and receive [training](#) specific to their job function.

50.2.6.4. Veterinarians will receive [training](#) specific to their job function.

50.2.6.5. Operators of veterinary units must work at the direction of a veterinarian and receive [training](#) specific to their job function.

50.2.6.6. Researchers using x-rays for non-human use in-vivo or in vitro will receive [training](#) specific to their job function.

50.3. *Can I Be In The Room During An X-Ray Procedure?*

Only personnel essential to the procedure should be in the exposure area during operation of the unit. All essential personnel should be adequately shielded during fluoroscopic procedures.

50.4. *Do I Have To Wear A Lead Apron?*

During radiographic procedures, if you are not able to stand behind an operator's barrier, then you must wear a protective, lead apron.

During fluoroscopic procedures, protective aprons shall be worn by the physician, nurse, technician, and all other persons within the room.

51. PATIENT SAFETY GUIDELINES**51.1. *Protocol Review***

Appropriate radiation delivered for the medical information acquired requires a commitment from practitioners, administration and medical directors.

51.1.1. All protocols should be reviewed for their potential to cause an injury to a patient.

51.1.2. Providers should modify protocols to optimize cumulative absorbed dose to any skin area.

51.2. *Fluoroscopy records*

51.2.1. Record radiation exposure information in patient's medical record for procedures involving fluoroscopy.

RAD-030, RADIATION SAFETY MANUAL

51.2.2. Identify areas of patient's skin that received the dose

51.2.3. Record total fluoroscopy time, and the typical technique (mA, kVp) generated during the procedure

51.2.4. Record cumulative air kerma and dose area product (DAP), if available

51.3. *Thresholds for follow up:*

51.3.1. Any procedure exceeding 5000 mGy air kerma, or 60 minutes fluoro time if air kerma is not available:

51.3.1.1. The radiologic technologists in Radiology and Cardiology techs for Cath lab and Electrophysiology will submit the completed [Skin Injury Dose Risk Estimate](#) Form to Radiation Safety (Appendix E).

51.3.1.2. The skin should be evaluated and its appearance documented in the patients' record. If warranted,

51.3.1.2.1. Attending, consulting dermatologist or consulting radiation oncologist should see the patient initially at 2 weeks and followed (by either the dermatologist or interventional proceduralist) for several months, as needed.

51.3.1.2.2. Documentation of the skin health shall be made in the medical record.

51.3.1.3. Assist Radiation Safety in evaluating the conditions of the exposure, the machine's performance, and any other criteria necessary for dose estimation.

51.3.1.4. File a SAFE report to follow the event.

51.3.2. Any procedures that are suspected to have resulted in a Joint Commission sentinel event exposure, >15 Gy peak skin dose, to a patient shall be reported to the Office of Quality by Radiation Safety.

52. USING X-RAYS IN RESEARCH

In order to obtain authorization to use x-rays on humans for research purposes (through the Institutional Review Board), see Human Use and Clinical Trials in ["Obtaining an Authorization" Section](#).

53. NON-HUMAN USE OF X-RAYS

Any research use of machine-produced radiation on animals or samples requires an application to the Radiation Safety Committee 2. A copy of this application can be found at the [X-Ray Application Non-Human Use form](#). General Guidelines for Safe Use of Non-Medical X-ray are available on the EHSO web page.

RAD-030, RADIATION SAFETY MANUAL
APPENDIX A - LABORATORY INSPECTION VIOLATIONS

Category	Item	Recommendation
Administrative Controls	4.4 Personnel working with radioactive materials are identified on PI's authorization permit.	Radioactive materials must only be handled by authorized and trained personnel.
Administrative Controls	4.6 The EHS Assist database reflects current inventory of radioactive materials stock vials, including record of volumes withdrawn from each stock vial.	Any quantity of radioactivity used must be recorded in EHS Assist. When a vial is empty, select 'totally disposed' to remove it from inventory.
Administrative Controls	4.7 The EHS Assist database reflects current inventory of radioactive waste containers, including record of activity discarded into each waste container.	Any quantity of radioactivity disposed of must be recorded in EHS Assist. When waste container is full, seal the container and request a waste pick-up.
Administrative Controls	4.9 Area Geiger meter surveys and wipe tests are performed during the work weeks that radioactive materials are used.	When radioactive materials are in use, Geiger meter surveys and wipe tests must be performed at least weekly in areas where radioactive material is used or stored.
Administrative Controls	4.10 Documentation of Geiger meter surveys includes the Geiger meter's model, serial number and calibration due date, date of the survey, and the initials of the person who performed the survey. The results are recorded in units of mR/hr and include a background reading.	Complete documentation on the appropriate forms when the Geiger meter survey is performed. Documentation demonstrates that the survey has occurred within the required timeframe.
Administrative Controls	4.11 Documentation of wipe tests include a list or map of areas surveyed, model and manufacturer of counter used, date of test, and the initials of the individual who performed the test. The results are either recorded in units of dpm or in cpm with counter efficiency and include a background reading.	Complete documentation on the appropriate forms when the wipe test is performed. Documentation demonstrates that the survey has occurred within the required timeframe.
Administrative Controls	4.12 If removable contamination is found, lab attempts decontamination of contaminated areas. Lab repeats the contamination survey and documents the clean-up effort.	Area is considered decontaminated when liquid scintillation counter results are less than 200 dpm /100 cm ² and Geiger counter readings are less than 2 mR/hr. If lab is unable to decontaminate area, contact your research radiation safety liaison for assistance.

RAD-030, RADIATION SAFETY MANUAL

General Radiation Safety	4.16 Use and storage of radioactive materials takes place in the authorized area.	Radioactive materials must be used and stored in authorized locations. Submit an amendment to add new locations.
Waste	4.24 Radioactive waste is segregated by isotope and waste type (Dry, Liquid, or Liquid Scintillation Vial).	Identify isotopes present in laboratory and ensure that each isotope has at least one waste container for dry, liquid and liquid scintillation vial waste. If lab does not use liquid scintillation vials except for wipe tests, lab may label liquid scintillation vial waste container as swipes only and does not have to distinguish between isotopes.
Waste	4.26 All radioactive trefoils on vials or other containers are defaced prior to disposal into the radioactive waste container.	Check all waste for trefoils prior to disposing in radiation waste containers.
Waste	4.28 Radioactive waste is not disposed of via sewer without authorization and documentation. Sewer disposal is not in excess of authorized limits.	Collect liquid radioactive waste in liquid radioactive waste containers unless specifically authorized to dispose of liquid radioactive waste via drain.